

1440

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

MAY 10 2016

FDLE

Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: KENNEDY SPACE CENTER  
Time of Inspection: 00:00

Date of Inspection: 05/05/2016

Serial Number: 80-001440  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:01316080A1 Exp: 02/05/2018
0.000	0.040 / 0.051	0.079	0.202	0.077
0.000	0.050 / 0.051	0.080	0.203	0.077
0.000	0.050 / 0.051	0.080	0.204	0.078

Number of Simulators Used: 5

Remarks:  
05: Control Outside Tolerance LOOSE CONNECTION.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



KAREN A JOYNER

Signature and Printed Name

05/05/2016  
Date

1440

RECEIVED

# Florida Department of Law Enforcement

APR 26 2016

## Alcohol Testing Program

FDLE  
Alcohol Testing Program

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: KENNEDY SPACE CENTER  
Time of Inspection: 20:12

Date of Inspection: 04/20/2016

Serial Number: 80-001440  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:01316080A1 Exp: 02/05/2018
0.000	0.051	0.081	0.203	0.078
0.000	0.052	0.081	0.203	0.079
0.000	0.051	0.081	0.204	0.079

Number of Simulators Used: 4

Remarks:

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

KAREN A JOYNER

Signature and Printed Name

04/20/2016  
Date

Florida Department of Law Enforcement MAR 09 2016  
Alcohol Testing Program FDLE  
Alcohol Testing Program  
AGENCY INSPECTION REPORT - INTOXILYZER 8000 BK

Agency: KENNEDY SPACE CENTER  
Time of Inspection: 08:27

Date of Inspection: 03/08/2016

Serial Number: 80-001440  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_


Remarks:

BYPASS TO OPERATE INSTRUMENT Non-compliance: .

*compliance not determined*

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JAKE L SHANAHAN

Signature and Printed Name

03/08/2016  
Date