

1441
SP

RECEIVED

NOV 22 2016

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: KENNEDY SPACE CENTER
Time of Inspection: 01:41

Date of Inspection: 11/17/2016

Serial Number: 80-001441
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:01316080A1 Exp: 02/05/2018
0.000	0.050	0.079	0.202	0.077
0.000	0.051	0.081	0.204	0.078
0.000	0.051	0.081	0.205	0.077

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



MAC W MC CARTHY

Signature and Printed Name

11/17/2016

1441 SP

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

NOV 03 2016

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: KENNEDY SPACE CENTER
Time of Inspection: 00:04

Date of Inspection: 10/31/2016

Serial Number: 80-001441
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:01316080A1 Exp: 02/05/2018
/ 0.000	0.050	0.079 / 0.080	0.202	0.078
/ 0.000	0.051	0.080 / 0.081	0.203	0.077
/ 0.000	0.051	0.080 / 0.080	0.205	0.078


Number of Simulators Used: 5

Remarks:

00: Ambient Fail POOR VENTILATION IN ROOM. 08: RFI Detect CELL PHONE IN HALLWAY.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 _____
Signature and Printed Name

MAC W MC CARTHY

10/31/2016
Date

1441

RECEIVED

SEP 30 2016

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: KENNEDY SPACE CENTER
Time of Inspection: 23:38

Date of Inspection: 09/22/2016

Serial Number: 80-001441
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201502G Exp: 02/24/2017	0.08g/210L Test (g/210L) Lot#: 201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 01316080A1 Exp: 02/05/2018
0.000	0.049	0.078	0.201	/ 0.077
0.000	0.049	0.078	0.202	/ 0.077
0.000	0.050	0.079	0.203	/ 0.076

Number of Simulators Used: 5

Remarks:
08: RFI Detect.

ENTERED WRONG INFO 0.05 SOLUTION LOT # 201507A, EXPIRY DATE 07/14/2017 ^{MAC}

MAC W MC
22 SEPT 14



The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MAC W MC Signature and Printed Name

MAC W MC CARTHY

09/22/2016
Date

Florida Department of Law Enforcement AUG 16 2016

Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: KENNEDY SPACE CENTER
Time of Inspection: 23:34

Date of Inspection: 08/11/2016

Serial Number: 80-001441
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:01316080A1 Exp: 02/05/2018
0.000	0.048	0.078	0.199	0.086 / 0.076
0.000	0.048	0.078	0.200	0.083 / 0.076
0.000	0.049	0.079	0.201	0.083 / 0.076

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MAC W MC CARTHY Signature and Printed Name

08/11/2016
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

MAY 24 2016

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Alcohol Testing Program

Agency: KENNEDY SPACE CENTER
Time of Inspection: 09:03

Date of Inspection: 05/23/2016

Serial Number: 80-001441
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____


QJAN

Number of Simulators Used: _____

Remarks:
BYPASS AI TO OPERATE INSTRUMENT Non-compliance: NOT DETERMINED.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JAKE L SHANAHAN

Signature and Printed Name

05/23/2016
Date

CCCCCCCCCCCCCCCC

144
RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

FEB 02 2016

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: KENNEDY SPACE CENTER
Time of Inspection: 23:24

Date of Inspection: 01/26/2016

Serial Number: 80-001441
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22315080A3 Exp: 09/05/2017
0.000	0.048	0.076	0.198	0.078
0.000	0.049	0.078	0.199	0.077
0.000	0.050	0.078	0.200	0.077

Number of Simulators Used: 4

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MAC W MC CARTHY
Signature and Printed Name

01/26/2016
Date