

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

OCT 27 2016

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BISCAYNE NAT PARK

Time of Inspection: 05:50 0955 (PE)

Date of Inspection: 10/26/2016 (PE)

Serial Number: 80-001442

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 22315080A2 Exp: 09/05/2017
0.000	0.049	0.078	0.199	0.081
0.000	0.049	0.077	0.199	0.081
0.000	0.048	0.078	0.200	0.081


Number of Simulators Used: 5

Remarks:

Agency Inspection conducted 10/26/2016, INSTRUMENT HAD
WRONG DATE/TIME

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


 _____ PABLO ESPINOZA
 Signature and Printed Name
 10/26/2016
~~10/01/2016~~ (PE)
 Date