## Florida Department of Law Enforcement DEC 16 2016 Alcohol Testing Program FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLAY COUNTY SO Time of Inspection: 16:55

Date of Inspection: 12/11/2016

Serial Number: 80-001443

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
	Yes	
Alcohol Free Subject Test: 0.000		
Mouth Alcohol Test: Slope Not Met	Yes	<b></b>
	Yes	
Interferent Detect Test: Interferent Detect		
Diagnostic Check (Post-Inspection): OK	Yes	
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:12716080A5 Exp: 06/05/2018
0.000	0.049	0.078	0.196	0.081
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.078	0.198	0.081

Munher	24	Simulators	Hond.	E

Remarks:

The above	instrument complie	es (X)	does not comply (	) with Chapter 11D-8, FAC.			
I certify performed	this inspection,	n accordance	with the provisions	Enforcement Agency Inspector s of Chapter 11D-8, FAC.	Permit	and that	1
demonstrate an acco	25%	Charu	Signature and Pr	CHARLES M HARWOOD			

12/11/2016 Date

### Florida Department of Law Enforcement Alcohol Testing Program

NOV 1 4 2016

FDLE Alcohol Testing Program

#### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLAY COUNTY SO Time of Inspection: 16:05

Date of Inspection: 11/06/2016

Serial Number: 80-001443

Software: 8100.27

Check or Test	YES	МО
Date and/or Time Adjusted	******	1 240
Diagnostic Charle (n		No
Diagnostic Check (Pre-Inspection): OK		
37 - 1 - 7	Yes	1
Alcohol Free Subject Test: 0.000		
	Yes	
Mouth Alcohol Test: Slope Not Met		
	Yes	
Interferent Detect Test: Interferent Detect		
	Yes	
Diagnostic Check (Post-Inspection): OK		
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:12716080A5 Exp: 06/05/2018
0.000	0.048	0.078	0.196	0.081
0.000	0.049	0.078	0.197	0.082
0.000	0.049	0.078	0.198	0.081

Number	of	Simulators	Heed.	Ę	

Remarks:

MONTHLY INSPECTION

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I this inspection in accordance with the provisions of Chapter 11D-8, FAC.

> RICHARD L PATRONE Signature and Printed Name

> > 11/06/2016 Date

# Florida Department of Law Enforcement of Alcohol Testing Program

UCI I 4 2016

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Testing Program

Agency: CLAY COUNTY SO Time of Inspection: 10:38

Date of Inspection: 10/09/2016

Serial Number: 80-001443 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	140	NO
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:12716080A5 Exp: 06/05/2018
0.000	0.049	0.077	0.196	0.080
0.000	0.049	0.077	0.196	0.080
0.000	0.049	0.078	0.197	0.080

Number of	S	imulators	Used:	5	
-----------	---	-----------	-------	---	--

Remarks:

MONTHLY INSPECTION

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DOPUTT K. GATRINE

7121,

RICHARD L PATRONE

Signature and Printed Name

### Florida Department of Law Enforcement Alcohol Testing Program

AUG 3 0 2016

FDLE
Alcohol Testing Program

#### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLAY COUNTY SO Time of Inspection: 17:21

Date of Inspection: 08/24/2016

Serial Number: 80-001443 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		+ -:-
		No
Diagnostic Check (Pre-Inspection): OK		
Algohol Prog Cubicat Waste O 000	Yes	
Alcohol Free Subject Test: 0.000	Ì	
Mouth Alcohol Test: Slope Not Met	Yes	
Model Alcohol lest: Slope Not Met		
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	ļ
	Yes	1

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:12716080A5 Exp: 06/05/2018
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.078	0.197	0.081
0.000	0.049	0.078	0.197	0.081

Number	οŕ	Simulators	Used:	5

Remarks:

The above	instrument complies ( X ) does not comply (	) with Chapter 11D-8, FAC.		
	that I hold a valid Florida Department of Law this inspection in accordance with the provision		Permit a	ind that I
	Signature and Pr.	CHARLES M HARWOOD		

08/24/2016 Date

## Florida Department of Law Enforcement SEP 2 8 2016 Alcohol Testing Program FDLE

Alcohol Testing Program

#### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLAY COUNTY SO Time of Inspection: 19:00

Date of Inspection: 09/25/2016

Serial Number: 80-001443 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		110
		No
Diagnostic Check (Pre-Inspection): OK		
Algobol Theo Cubicat Tart 0 000	Yes	
Alcohol Free Subject Test: 0.000	L. Constanting of the Constantin	
Mouth Alcohol Test: Slope Not Met	Yes	
Modern Alcohol lest. Slope Rot Met		
Interferent Detect Test: Interferent Detect	Yes	
The second secon	Yes	
Diagnostic Check (Post-Inspection): OK	ies	
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:12716080A5 Exp: 06/05/2018
0.000	0.048	0.078	0.196	0.080
0.000	0.048	0.079	0.197	0.080
0.000	0.048	0.079	0.197	0.080

Number of	Simulators	Used:	5
-----------	------------	-------	---

Remarks:

MONTHLY INSPECTION

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

\_\_\_\_\_\_

RICHARD L PATRONE

09/25/2016 Date TOTTOM Department of Law Emolecement

## Alcohol Testing Program

80-001443

#### AGENCY INSPECTION PROCEDURES - INTOXILYZER 8000

Prepare at least two simulators for use and allow them to warm up for at least thirty minutes prior to the first analysis. When changing solutions, allow the new solution to warm up for at least ten minutes after the heater light turns off for the first time. Ensure that each simulator maintains an air leak resistant seal and an operational temperature of 34 C (+/- 0.2 C).

Only distilled or deionized water must be used for the Alcohol Free Test and the Interferent Detect Test. Class A glassware must be used when measuring solutions.

Only approved and non-expired alcohol reference solution and non-expired dry gas standard from an approved source must be used during the applicable portions of the inspection.

4. The instrument must display READY MODE prior to beginning the inspection.

Press ESC twice to access the main menu. Enter Agency inspector last name, first name and middle initial at USER prompts. Press the 2 key to access the Agency Inspector menu. Enter password. Scroll to Inspection Test and press ENTER. Enter Agency Inspector last name, first name and middle initial at INSPECTOR prompts.

All results are reported to three decimal places in g/210L. Results of all air blanks must display 0.000. The instrument will abort the inspection process if the air blank result is not 0.000. If any check or test is out of compliance, the instrument will prompt the Agency Inspector to REPEAT (Y/N) the check or test. Each check or test may only be repeated once. If a check or test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report – Intoxilyzer 8000.

Check DATE. Adjust if necessary. Press ENTER. Check TIME. Adjust if necessary. Press ENTER.

DIAGNOSTIC CHECK (Pre-Inspection). Press ENTER. All checks must display OK. If any check is not OK, the instrument will abort the inspection process.

NUMBER OF SIMULATORS USED. Enter the number of simulators used during the inspection.

ALCOHOL FREE SUBJECT/MOUTH ALCOHOL TEST. Press ENTER. When PROVIDE SAMPLE NOW is displayed, introduce an alcohol-free breath sample into the instrument. The instrument must display a result of 0.000. Rinse mouth with mouth alcohol solution. When PROVIDE SAMPLE NOW is again displayed introduce a breath sample into the instrument. The instrument must display SLOPE NOT MET.

ALCOHOL FREE TEST. Attach a simulator containing 500 mL distilled or deionized water to the instrument. Press ENTER. Conduct three (3) analyses. The results of all three (3) analyses must be displayed as 0.000.

INTERFERENT DETECT TEST. Attach a simulator containing 3 mL of acetone stock solution and 500 mL distilled or deionized water to the instrument. Press ENTER. Conduct three (3) analyses. The instrument must display INTERFERENT DETECT for each analysis.

0.05 g/210L TEST. Attach a simulator containing 0.05 g/210L alcohol reference solution to the instrument. Press ENTER. Enter the lot number and expiration date of the alcohol reference solution used. Conduct three (3) analyses. The displayed results of all three (3) analyses must be within the acceptable range.

0.08 g/210L TEST. Attach a simulator containing 0.08 g/210L alcohol reference solution to the instrument. Press ENTER. Enter the lot number and expiration date of the alcohol reference solution used. Conduct three (3) analyses. The displayed results of all three (3) analyses must be within the acceptable range.

0.20 g210L TEST. Attach a simulator containing 0.20 g/210L alcohol reference solution to the instrument. Press ENTER. Enter the lot number and expiration date of the alcohol reference solution used. Conduct three (3) analyses. The displayed results of all three (3) analyses must be within the acceptable range.

0.08 g/210L DRY GAS STANDARD TEST. Attach a cylinder containing 0.08 g/210L dry gas standard to the instrument. Press ENTER. Enter the lot number and expiration date of the dry gas standard used. Conduct three (3) analyses. The displayed results of all three (3) analyses must be within the acceptable range.

DIAGNOSTIC CHECK (Post-Inspection). Press ENTER. All checks must display OK. If any check is not OK, the instrument will abort the inspection process.

REVIEW REMARKS. Enter Y/N.

IN COMPLIANCE. Enter Y/N to state whether the instrument complies or does not comply with the requirements of Chapter 11D-8, FAC. If the instrument does not comply with Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector.

The results of the Agency Inspection must be recorded on FDLE/ATP Form 40 Agency Inspection Report – Intoxilyzer 8000. For regulatory and administrative purposes only, the results of the agency inspection must be made electronically available to the Department within five (5) business days of completing the inspection.

RECEIVED

AUG 3 0 2016

FDLE Alcohol Testing Program 4-16 St. Spussed

FDLE/ATP Form 39 - March 2004, Ref. 11D-8.006

## Florida Department of Law Enforcement UL 0 8 2016 Alcohol Testing Program FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLAY COUNTY SO Time of Inspection: 15:04

Date of Inspection: 07/03/2016

Serial Number: 80-001443 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
		No
Diagnostic Check (Pre-Inspection): OK		
51 5 7	Yes	
Alcohol Free Subject Test: 0.000		
52	Yes	
Mouth Alcohol Test: Slope Not Met		
	Yes	
Interferent Detect Test: Interferent Detect		
	Yes	
Diagnostic Check (Post-Inspection): OK		
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:12716080A5 Exp: 06/05/2018
0.000	0.049	0.078	0.196	0.079
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.078	0.198	0.078

Number	ο£	Simulators	Used:	5

Remarks:

MONTELY INSPECTION

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed that inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

07/03/2016 Date

## JUL 0 8 2016

#### Florida Department of Law Enforcement **Alcohol Testing Program** Alcohol Testing Program

#### AGENCY INSPECTION REPORT - INTOXILYZER

Agency: CLAY COUNTY SO Time of Inspection: 14:23

Date of Inspection: 06/11/2016

Serial Number: 80-001443 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
		No
Diagnostic Check (Pre-Inspection): OK		
33 - 1 - 7 - 77	Yes	
Alcohol Free Subject Test: 0.000		
Mouth Alcohol Test: Slope Not Met	Yes	
worth riconor test: Stobe not Wet		
Interferent Detect Test: Interferent Detect	Yes	
	***	
Diagnostic Check (Post-Inspection): OK	Yes	<del> </del>
-1331000110 Officer (1050-185peccion): Or	•••	
	Yes	ì

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A3 Exp: 10/05/2017
0.000	0.049	0.078	0.201	0.079
0.000	0.049	0.078	0.200	0.079
0.000	0.049	0.078	0.201	0.079

Number	of	Simulators	Used:	5	
--------	----	------------	-------	---	--

Remarks:

I blocked the mouth piece while placing the brenth tube in the cradle. algors A F / M A: Improper Sample.

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 110-8, FAC.

Signature and Printed Name

CHARLES M HARWOOD

06/11/2016

## Florida Department of Law Enforcement Alcohol Testing Program

MAY 0 2 2016

FDLE

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Testing Program

Agency: CLAY COUNTY SO Time of Inspection: 14:24

Date of Inspection: 04/27/2016

Serial Number: 80-001443 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
braghostic check (Fre-Inspection): OK		
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
-	Yes	
Interferent Detect Test: Interferent Detect		<del>                                     </del>
Diagnostic Check (Post-Inspection): OK	Yes	
orangeoutic check (FOSC-Inspection): OK		
	Yes	1

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A3 Exp: 10/05/2017
0.000	0.048	0.077	0.197	0.080
0.000	0.048	0.077	0.199	0.080
0.000	0.048	0.077	0.199	0.080

Number	οf	Simulators	Useđ:	5	

Remarks:

MONTHLY INSPECTION RETURN FOR ANNUAL

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

RICHARD L PATRONE Signature and Printed Name

04/27/2016 Date

## Florida Department of Law Enforcement MAR 2 5 2016 Alcohol Testing Program FDLE

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLAY COUNTY SO Time of Inspection: 16:56

Date of Inspection: 03/21/2016

Serial Number: 80-001443 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	4.213	NO.
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		
	Yes	}

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A3 Exp: 10/05/2017
0.000	0.049	0.079	0.200	0.082
0.000	0.050	0.079	0.201	0.082
0.000	0.050	0.079	0.201	0.081

Number of Simulators Used: 5

Remarks:

MONTHLY INSPECTION

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this Aspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

RICHARD L PATRONE

03/21/2016

Date

## Florida Department of Law Enforcement Alcohol Testing Program Alcohol Testing Program

FDLE Alcohol Testing Program

#### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLAY COUNTY SO Time of Inspection: 18:11

Date of Inspection: 02/21/2016

Serial Number: 80-001443 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		1.0
		No
Diagnostic Check (Pre-Inspection): OK		
	Yes	
Alcohol Free Subject Test: 0.000		
	Yes	
Mouth Alcohol Test: Slope Not Met		
	Yes	
Interferent Detect Test: Interferent Detect		
	Yes	
Diagnostic Check (Post-Inspection): OK		
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A3 Exp: 10/05/2017
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.077	0.200	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

CHAR

Signature and Printed Name

02/21/2016 Date

## » RECEIVED

## Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

#### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLAY COUNTY SO Time of Inspection: 20:46

Date of Inspection: 01/21/2016

Serial Number: 80-001443

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
		No
Diagnostic Check (Pre-Inspection): OK		
	Yes	
Alcohol Free Subject Test: 0.000		
	Yes	
Mouth Alcohol Test: Slope Not Met		
	Yes	
Interferent Detect Test: Interferent Detect		1
	Yes	
Diagnostic Check (Post-Inspection): OK		
	Yes	-

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A3 Exp: 10/05/2017
0.000	0.050	0.079	0.201	0.080
0.000	0.050	0.079	0.201	0.080
0.000	0.050	0.079	0.202	0.080

Number	of	Simulators	Used:	5

Remarks:

MONTHLY INSPECTION

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

724

RICHARD L PATRONE

Signature and Printed Name

01/21/2016 Date