

Florida Department of Law Enforcement Alcohol Testing Program

17W
RECEIVED
OCT 03 2016

AGENCY INSPECTION REPORT - INTOXILYZER 8000 FDLE Alcohol Testing Program

Agency: OSCEOLA COUNTY S.O.
Time of Inspection: 09:20

Date of Inspection: 09/29/2016

Serial Number: 80-001715
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1589901 Exp: 10/30/2016
0.000	0.049	0.078	0.199	0.081
0.000	0.050	0.079	0.201	0.081
0.000	0.049	0.079	0.201	0.081

Number of Simulators Used: 5

Remarks:



The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

ANGEL M DIAZ

Signature and Printed Name

09/29/2016
Date