

RECEIVED

SEP 28 2016

1721

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BAY P.D.

Serial Number: 80-001721

Time of Inspection: 11:32

Date of Inspection: 09/25/2016

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A2 Exp: 10/05/2017
0.000	0.050	0.081	0.200	0.078
0.000	0.051	0.080	0.199	0.078
0.000	0.050	0.080	0.200	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

EDWIN J LUTZ

Signature and Printed Name

09/25/2016
Date

1266

RECEIVED

SEP 28 2016

Florida Department of Law Enforcement

FDLE

Alcohol Testing Program

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BAY P.D.
Time of Inspection: 11:23

Date of Inspection: 09/25/2016

Serial Number: 80-001266
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A2 Exp: 10/05/2017
0.000	0.048	0.079	0.200	0.080
0.000	0.048	0.079	0.202	0.079
0.000	0.049	0.079	0.201	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

EDWIN J LUTZ

Signature and Printed Name

09/25/2016
Date

Florida Department of Law Enforcement Alcohol Testing Program

1721

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BAY P.D.
Time of Inspection: 10:37

Date of Inspection: 08/23/2016

Serial Number: 80-001721
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A2 Exp: 10/05/2017
0.000	0.050	0.080	0.198	0.078
0.000	0.050	0.081	0.198	0.078
0.000	0.050	0.081	0.199	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

EDWIN J LUTZ
Signature and Printed Name

08/23/2016
Date

1721
D

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BAY P.D.

Serial Number: 80-001721

Time of Inspection: 11:48

Date of Inspection: 05/23/2016

Software: 8100.27

Check or Test	YES	NO
Flow and/or Time Adjusted		No
Manufacturer Check (Pre-Inspection): OK	Yes	
Flow and/or Time Adjusted Test: 0.000	Yes	
Check Alcohol Test: Slope Not Met	Yes	
Interferent Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Conc Std (g/210L)	0.05g/210L Test (g/210L) Lot#: 2514015 Exp: 02/24/2017	0.08g/210L Test (g/210L) Lot#: 2014005 Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 7014012 Exp: 02/24/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 241500072 Exp: 02/24/2017
0.000	0.050	0.080	0.198	0.072
0.050	0.051	0.081	0.199	0.073
0.100	0.051	0.081	0.199	0.073

Inspected by: [Signature]

This instrument complies (X) does not comply () with Chapter 112-8, F.S.

Signature and Printed Name

Date

Dan Morgan [Signature]