

*Amended*

*1743*

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

NOV 09 2016

FDLE

Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: JACKSONVILLE SO  
Time of Inspection: 09:59

Date of Inspection: 07/25/2016

Serial Number: 80-001743  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:01316080A2 Exp: 02/05/2016 - 2018 <i>RST</i>
0.000	0.049	0.081	0.208	0.081
0.000	0.049	0.081	0.208	0.081
0.000	0.050	0.080	0.208	0.081

Number of Simulators Used: 5

Remarks:

*wrong date on .080 Dry Gas should  
be exp. date 02.05.2018*

*RST 7326*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Robert D Thomason 7326*

ROBERT D THOMASON  
Signature and Printed Name

07/25/2016  
Date