

80-001865 RWS

RECEIVED  
DEC 06 2016

# Florida Department of Law Enforcement Alcohol Testing Program

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD  
Time of Inspection: 10:21

Date of Inspection: 12/06/2016

Serial Number: 80-001865  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22315080A2 Exp: 09/05/2017
0.000	0.048	0.077	0.195	0.078
0.000	0.048	0.078	0.196	0.078
0.000	0.049	0.079	0.197	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



GORDON L FRANZ  
Signature and Printed Name

12/06/2016  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**  
OCT 11 2016  
FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD  
Time of Inspection: 12:38

Date of Inspection: 09/26/2016

Serial Number: 80-001865  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.045	0.000 / 0.067		
0.000	0.046	0.000 / 0.071		
0.000	0.046	0.000 / 0.072		

Number of Simulators Used: 5

**Remarks:**

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Gordon L. Franz*

GORDON L. FRANZ

Signature and Printed Name

09/26/2016  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

AUG 22 2016

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD  
Time of Inspection: 09:18

Date of Inspection: 08/22/2016

Serial Number: 80-001865  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 22315080A2 Exp: 09/05/2017
0.000	0.047	0.077	0.193	0.078
0.000	0.047	0.077	0.195	0.078
0.000	0.048	0.078	0.194	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Gordon L. Franz*

GORDON L. FRANZ

Signature and Printed Name

08/22/2016  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**

OCT 19 2016

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD  
Time of Inspection: 12:38

Date of Inspection: 09/26/2016

Serial Number: 80-001865  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____
0.000	0.045	0.000 / 0.067		
0.000	0.046	0.000 / 0.071		
0.000	0.046	0.000 / 0.072		

Number of Simulators Used: 5

**Remarks:**

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

updated 10/18/16: AFTER INITIAL .08 INSPECTION, SIMULATOR WAS CLEANED, REPLENISHED WITH FRESH .08 ALCOHOL REFERENCE SOLUTION, ALLOWED TO WARM UP AND SUBSEQUENTLY RETESTED. 2ND TEST INDICATED LOW READINGS. *G. Franz*

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Gordon L. Franz*

GORDON L FRANZ

Signature and Printed Name

09/26/2016  
Date

80-001865-003

# Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED  
JUL 25 2016  
FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD  
Time of Inspection: 09:12

Date of Inspection: 07/25/2016

Serial Number: 80-001865  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22315080A2 Exp: 09/05/2017
0.000	0.047	0.076	0.194	0.079
0.000	0.048	0.077	0.194	0.079
0.000	0.048	0.077	0.194	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Gordon L. Franz  
Signature and Printed Name

GORDON L FRANZ

07/25/2016  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**

JUN 27 2016

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD  
Time of Inspection: 11:44

Date of Inspection: 06/13/2016

Serial Number: 80-001865  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22315080A2 Exp: 09/05/2017
0.000	0.048	0.077	0.195	0.079
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.078	0.196	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Gordon L. Franz*

GORDON L FRANZ

Signature and Printed Name

06/13/2016  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**  
JUN 06 2016  
FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD  
Time of Inspection: 12:49

Date of Inspection: 06/02/2016

Serial Number: 80-001865  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_

Remarks:  
BYPASSED AI TO OPERATE INSTRUMENT. NOT A COMPLIANCE CHECK.

*PMW*

*N/A RMB*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*[Signature]*

DANIELLE M BELL

Signature and Printed Name

06/02/2016  
Date