

# Florida Department of Law Enforcement Alcohol Testing Program

*15K*

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PORT ST LUCIE PD  
Time of Inspection: 11:04

Date of Inspection: 10/06/2016

Serial Number: 80-001962  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_

Remarks:  
BYPASSED AI TO OPERATE INSTRUMENT

*PLB*

N/A COMPLIANCE NOT DETERMINED

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Shayla Platt*

SHAYLA D PLATT

Signature and Printed Name

10/06/2016  
Date

Instrument # 80-001962

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PORT ST LUCIE PD  
Time of Inspection: 11:32

Date of Inspection: 08/29/2016

Serial Number: 80-001962  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:24515080A3 Exp: 10/05/2017
0.000	0.046	0.076	0.194	0.074 / 0.075
0.000	0.047	0.077	0.196	0.074 / 0.075
0.000	0.048	0.078	0.197	0.073 / 0.074

Number of Simulators Used: 4

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: **CHECKED SEALS, RETRIED TEST, STILL OUTSIDE TOLERANCE. UNKNOWN REASON. 2/29/16**

RECEIVED

OCT 09 2016

FDLE  
Alcohol Testing Program

SP

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

IAN HARRIS

Signature and Printed Name

08/29/2016  
Date