Florida Department of Law Enforcement Alcohol Testing Program

DEC 14 2016

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: NAS KEY WEST Time of Inspection: 08:19

Date of Inspection: 12/07/2016

Serial Number: 80-002170

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		
		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

	0 000
1	ACCU

Number of Simulators Used:

Remarks:

COMPLIANCE NOT DETERMINED, AI NOT CONDUCTED, DISSABLED MODE

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Elegn Rup Ruitu
Signature and Printed Name

DAVID E REYES-RIVERA

12/07/2016 Date

Florida Department of Law Enforcement Alcohol Testing Program

SEP 06 2016

FOLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: NAS KEY WEST

Time of Inspection: 17:40

Date of Inspection: 06/07/2016

Serial Number: 80-002170

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
		No
Diagnostic Check (Pre-Inspection): OK		1
	Yes	
Alcohol Free Subject Test: 0.000		
	Yes	1
Mouth Alcohol Test: Slope Not Met		
	Yes	
Interferent Detect Test: Interferent Detect	*	
	Yes	
Diagnostic Check (Post-Inspection): OK	-	
,	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:366849 Exp: 03/30/2019
0.000	0.047	0.079	0.195	0.081
0.000	0.047	0.079	0.196	0.081
0.000	0.047	0.079	0.196	0.081

Number of Simulators Used: 5____

Remarks: INSTRUMENT DID NOT PASS INITIAL TEST DUE TO OPERATER
INSTRUMENT DID NOT PASS INITIAL TEST DUE TO OPERATER
INSPECTION COMPLETED WITH PROPER DATA VERTEYING INSTRUMENT IN COMPLEME.
INSPECTION COMPLETED WITH PROPER DATA VERTEYING INSTRUMENT IN COMPLEME.

The above instrument complies (X) does not comply () with Chapter 11p-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

LEE B ANDERSON

06/07/2016

Date

FDLE/ATP Form 40 - March 2004





AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

JUL **06** 2016 FDLE

Alcohol Testing Program

Agency: Naval Air Station Key West Instrument Serial Number: 80-002170 AGENCY INSPECTION DATA REVIEW Agency Inspector: LT. Lee Anderson Date of Inspection: 06/07/2016 Time of Inspection: 16:55:31 Agency Inspection Discrepancy: Incomplete Untimely/Not Received Erroneous Information Procedural X Other Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded. Lot Number ☐ Expiration Date for _g/210L □Alcohol Reference Solution □Dry Gas Standard is □Incorrect □Expired. FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report - Intoxilyzer 8000. The REASON for repeating the following test(s); OR the Possible Cause and Corrective Action Taken on the following test(s) was not recorded: Alcohol Free Subject Test Mouth Alcohol Test ☐ Alcohol Free Test Interferent Detect Test 0.05 g/210L Test 0.08 g/210L Test 0.20 g/210L Test П X 0.08 g/210L Dry Gas Standard Test FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector." The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC. The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use. The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use. The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC. Other: OTHER ELECTRONIC DATA REVIEW **Loain Records** Comments: Date: Cylinder Change Records Date: Control Test Records Date: Diagnostic Check Records **CORRECTIVE ACTION** Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by 8/6/2016 (Date). Provide a written explanation regarding the referenced item(s) to the Department Inspector by (Date). Upload the Agency Inspection(s). Remove the instrument from evidentiary use until otherwise directed by the Department. No action required Other:

> 7/7/2016 Date

Signature of Alcohol Testing Program Staff Member