

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
DEC 01 2016
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO
Time of Inspection: 12:06

Date of Inspection: 12/01/2016

Serial Number: 80-003347
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:03415080A1 Exp: 03/05/2017
0.000	0.049	0.077	0.200	0.076
0.000	0.050	0.078	0.202	0.077
0.000	0.050	0.078	0.202	0.076

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Ricky A Frisco 3184 _____ RICKY A FRISCO
Signature and Printed Name

12/01/2016
Date

RECEIVED

NOV 03 2016

Florida Department of Law Enforcement
Alcohol Testing Program

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO
Time of Inspection: 09:26

Date of Inspection: 11/03/2016

Serial Number: 80-003347
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:03415080A1 Exp: 03/05/2017
0.000	0.049	0.078	0.201	0.076
0.000	0.049	0.079	0.202	0.076
0.000	0.050	0.079	0.202	0.076

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


3184
RICKY A FRISCO

Signature and Printed Name

11/03/2016
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
OCT 14 2016
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO
Time of Inspection: 08:47

Date of Inspection: 10/14/2016

Serial Number: 80-003347
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:03415080A1 Exp: 03/05/2017
0.000	0.050	0.078	0.200	0.077
0.000	0.050	0.078	0.201	0.076
0.000	0.050	0.078	0.201	0.076

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



RICKY A FRISCO

Signature and Printed Name

3184

10/14/2016
Date

RECEIVED

AUG 19 2016

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

 Agency: PASCO COUNTY SO
 Time of Inspection: 11:06

Date of Inspection: 08/19/2016

 Serial Number: 80-003347
 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:03415080A1 Exp: 03/05/2017
0.000	0.050	0.077	0.201	0.077
0.000	0.050	0.078	0.201	0.077
0.000	0.050	0.079	0.202	0.078

Number of Simulators Used: 5

Remarks:

 The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


3184
Signature and Printed Name

RICKY A FRISCO

08/19/2016
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

APR 08 2016

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO
Time of Inspection: 12:09

Date of Inspection: 04/08/2016

Serial Number: 80-003347
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:03415080A1 Exp: 03/05/2017
0.000	0.049	0.079	0.200	0.076
0.000	0.049	0.078	0.201	0.077
0.000	0.050	0.078	0.201	0.077

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Ricky A. Frisco 3184 RICKY A FRISCO
Signature and Printed Name

04/08/2016
Date

RECEIVED

Florida Department of Law Enforcement MAR 16 2016 Alcohol Testing Program

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO
Time of Inspection: 10:15

Date of Inspection: 03/16/2016

Serial Number: 80-003347
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:03415080A1 Exp: 03/05/2017
0.000	0.048	0.078	0.201	0.077
0.000	0.048	0.079	0.202	0.077
0.000	0.049	0.079	0.202	0.076

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Ricky A Frisco
 3184
 Signature and Printed Name RICKY A FRISCO

03/16/2016
Date

RECEIVED

FEB 29 2016

FDLE
Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO
Time of Inspection: 20:55

Date of Inspection: 02/25/2016

Serial Number: 80-003347
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 03415080A1 Exp: 03/05/2017
0.000	0.050	0.077	0.200	0.079
0.000	0.050	0.079	0.201	0.079
0.000	0.050	0.079	0.201	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Ricky A Frisco 3184 RICKY A FRISCO
Signature and Printed Name

02/25/2016
Date

RECEIVED

MAR 10 2016

FDLE Alcohol Testing Program

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Pasco County S.O.

Serial Number: 80-003347

Time of Inspection: 2000 Date of Inspection: 02/25/2016

Software: 8100.27

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		NO
Diagnostic Check (Pre-Inspection): OK	yes	
Alcohol Free Subject Test: 0.000	yes	
Mouth Alcohol Test: Slope Not Met	yes	
Interferent Detect Test: Interferent Detect	yes	
Diagnostic Check (Post-Inspection): OK		NO

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.00				
0.00				
0.00				

Number of Simulators Used: 5

Remarks:
 (28) ON 2/25/16 @ Approx. 2000 HRS. I STARTED AN AGENCY INSPECTION ON THIS INSTRUMENT. DURING THE .05 TEST I BUMPED THE POWER SWITCH TO THE POWER STRIP IT WAS PLUGGED INTO WHILE MOVING A SIMULATOR WHICH CAUSED THE INSTRUMENT TO POWER OFF. PRIOR TO THIS HAPPENING ALL CHECKS/TESTS PASSED AND WERE WITHIN TOLERANCE. I RE-BOOTED THE INSTRUMENT AND COMPLETED A SUCCESSFUL AGENCY INSPECTION FROM THE BEGINNING.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Ricky A. Frasco
3184

Ricky A. Frasco

02/25/16
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

RECEIVED
JUL 07 2016
FDLE
Alcohol Testing Program

Agency: PASCO COUNTY SO
Time of Inspection: 10:04

Date of Inspection: 07/07/2016

Serial Number: 80-003347
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:03415080A1 Exp: 03/05/2017
0.000	0.049	0.079	0.201	0.077
0.000	0.050	0.078	0.201	0.077
0.000	0.050	0.079	0.202	0.077

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Ricky A Frisco 3184 _____ RICKY A FRISCO
Signature and Printed Name

07/07/2016
Date

80-003347 *RWB*

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
JUN 03 2016
FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO
Time of Inspection: 09:46

Date of Inspection: 06/03/2016

Serial Number: 80-003347
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:03415080A1 Exp: 03/05/2017
0.000	0.050	0.079	0.202	0.077
0.000	0.051	0.080	0.203	0.077
0.000	0.051	0.080	0.202	0.077

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Ricky A. Frisco 3184 RICKY A FRISCO
Signature and Printed Name

06/03/2016
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

RECEIVED

MAY 05 2016

FDLE
Alcohol Testing Program

Agency: PASCO COUNTY SO
Time of Inspection: 09:40

Date of Inspection: 05/05/2016

Serial Number: 80-003347
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:03415080A1 Exp: 03/05/2017
0.000	0.050	0.078	0.201	0.077
0.000	0.050	0.079	0.202	0.077
0.000	0.050	0.079	0.202	0.077

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Ricky A Frisco
3184

RICKY A FRISCO

Signature and Printed Name

05/05/2016
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

MAR 10 2016

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Pasco County S.O.Serial Number: 80-003347Time of Inspection: 2000 Date of Inspection: 02/25/2016Software: 8100.27

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		NO
Diagnostic Check (Pre-Inspection): OK	yes	
Alcohol Free Subject Test: 0.000	yes	
Mouth Alcohol Test: Slope Not Met	yes	
Interferent Detect Test: Interferent Detect	yes	
Diagnostic Check (Post-Inspection): OK		NO

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.00				
0.00				
0.00				

Number of Simulators Used: 5

Remarks:

(25) ON 2/25/16 @ APPROX. 2000 HRS. I STARTED AN AGENCY INSPECTION ON THIS INSTRUMENT. DURING THE .05 TEST I BUMPED THE POWER SWITCH TO THE POWER STRIP IT WAS PLUGGED INTO WHILE MOVING A SIMULATOR WHICH CAUSED THE INSTRUMENT TO POWER OFF. PRIOR TO THIS HAPPENING ALL CHECKS/TESTS PASSED AND WERE WITHIN TOLERANCE. I RE-BOOTED THE INSTRUMENT AND COMPLETED A SUCCESSFUL AGENCY INSPECTION FROM THE BEGINNING.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Ricky A. Frasco
3184

Signature and Printed Name

Ricky A. Frasco

02/25/16
Date

DRAFT

Florida Department of Law Enforcement Alcohol Testing Program

MAR 02 2016

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY S.O.

Serial Number: 80-003347

Time of Inspection: 2000 Date of Inspection: 02/25/2016 Software: 8100.27

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		NO
Diagnostic Check (Pre-Inspection): OK	yes	
Alcohol Free Subject Test: 0.000	yes	
Mouth Alcohol Test: Slope Not Met	yes	
Interferent Detect Test: Interferent Detect	yes	
Diagnostic Check (Post-Inspection): OK	yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: 5

Remarks:

PP ON 2/25/16 @ Apprx. 2000 HRS. I STARTED AN AGENCY INSPECTION ON THIS INSTRUMENT. DURING THE .05 TEST I BUMPED THE POWER SWITCH TO THE POWER STRIP IT WAS PLUGGED INTO WHILE MOVING A SIMULATOR WHICH CAUSED THE INSTRUMENT TO POWER OFF. PRIOR TO THIS HAPPENING ALL CHECKS/TESTS PASSED AND WERE WITHIN TOLERANCE. I RE-BOOTED THE INSTRUMENT AND COMPLETED A SUCCESSFUL AGENCY INSPECTION FROM THE BEGINNING. *Ricky A. Frisco*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Ricky A. Frisco
3184

Ricky A. Frisco
Signature and Printed Name

02/25/16
Date

DRAFT

RECEIVED

FEB 29 2016

FDLE
Alcohol Testing

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO
Time of Inspection: 19:29

Date of Inspection: 02/25/2016

Serial Number: 80-003347
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: 5

Remarks:

A F / M A: PROVIDED 2 ALC FREE SAMPLES Non-compliance: PROVIDED 2 ALC FREE SAMPLES.

2/25/16

(PF)

I PROVIDED (2) ALCOHOL FREE BREATH SAMPLES TO START IN FERROP. STARTED NEW AGENCY INSPECTION FROM THE BEGINNING AND EVERYTHING WAS GOOD.

Ricky A Frisco - 3184

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Ricky A Frisco 3184 RICKY A FRISCO
Signature and Printed Name

02/25/2016
Date