

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

JUN 01 2016

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 11:36

Date of Inspection: 06/01/2016

Serial Number: 80-003407
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28814080A3 Exp: 11/05/2016
0.000	0.049	0.079	0.197	0.079
0.000	0.050	0.079	0.197	0.079
0.000	0.050	0.080	0.197	0.079

Number of Simulators Used: 5

Remarks:
NA

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Susan M Barge

SUSAN M BARGE

Signature and Printed Name

06/01/2016
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

MAY 03 2016

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 15:39

Date of Inspection: 05/03/2016

Serial Number: 80-003407
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201502G Exp: 02/23/2017	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 28814080A3 Exp: 11/05/2016
0.000	0.050	0.079	0.200	0.078
0.000	0.051	0.079	0.200	0.077
0.000	0.050	0.079	0.200	0.077

Number of Simulators Used: 5

Remarks:
NONE

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



SUSAN M BARGE

Signature and Printed Name

05/03/2016
Date

RECEIVED

Florida Department of Law Enforcement

Alcohol Testing Program

APR 06 2016

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 10:04

Date of Inspection: 04/06/2016

Serial Number: 80-003407
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:26814080AD Exp: 11/05/2016
0.000	0.052	0.080	0.201	0.079
0.000	0.051	0.081	0.200	0.080
0.000	0.051	0.081	0.201	0.079

Number of Simulators Used: 5

Remarks:
NA

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Susan N Barge

SUSAN N BARGE

Signature and Printed Name

04/06/2016
Date

8880-003407

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

FEB 09 2016

AGENCY INSPECTION REPORT - INTOXILYZER 8000

FDLE
Alcohol Testing Program

Agency: FL HIGHWAY PATROL
Time of Inspection: 15:28

Date of Inspection: 02/01/2016

Serial Number: 80-003407
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 26814080A Exp: 07/04/2017	0.08g/210L Test (g/210L) Lot#: 26814080B Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 26814080C Exp: 05/17/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 26814080A3 Exp: 11/08/2016
0.000	0.050	0.079	0.198	0.080
0.000	0.050	0.079	0.199	0.080
0.000	0.050	0.079	0.199	0.080

Number of Simulators Used: 0

Remarks:
None

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

GREGORY J. HANLEY

02/01/2016
Date

2880-003407

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
JAN 29 2016

AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

FDLE

Agency: FL HIGHWAY PATROL
Time of Inspection: 16:33

Date of Inspection: 01/28/2016

Serial Number: 80-003407
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 02/24/2017	0.08g/210L Test (g/210L) Lot#:201502C Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28814080A3 Exp: 11/05/2016
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.080	0.199	0.078
0.000	0.049	0.080	0.199	0.078

Number of Simulators Used: 5

Remarks:
NONE

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Gregory J. Lamont

Signature and Printed Name

01/28/2016
Date