

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**  
JUN 01 2016  
FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL  
Time of Inspection: 11:39

Date of Inspection: 06/01/2016

Serial Number: 80-003412  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 16014080A2 Exp: 07/05/2016
0.000	0.050	0.079	0.196	0.078
0.000	0.050	0.078	0.197	0.079
0.000	0.050	0.078	0.197	0.078

Number of Simulators Used: 5

Remarks:  
NA

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Susan M Barge*

SUSAN M BARGE

Signature and Printed Name

06/01/2016  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

MAY 03 2016

FDLE  
Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL  
Time of Inspection: 15:47

Date of Inspection: 05/03/2016

Serial Number: 80-003412  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201502G Exp: 03/24/2017	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 16014880A2 Exp: 07/05/2016
0.000	0.050	0.079	0.199	0.078
0.000	0.050	0.079	0.198	0.078
0.000	0.050	0.079	0.198	0.078

Number of Simulators Used: 5

Remarks:  
NONE

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Cpl Susan M Barge*

SUSAN M BARGE

Signature and Printed Name

05/03/2016  
Date

RECEIVED

APR 06 2016

FDLE  
Alcohol Testing Program

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL  
Time of Inspection: 10:53

Date of Inspection: 04/06/2016

Serial Number: 80-003412  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201500A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#: 201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 16014088A2 Exp: 07/05/2016
0.000	0.050	0.079	0.200	0.079
0.000	0.051	0.080	0.200	0.079
0.000	0.050	0.080	0.200	0.079

Number of Simulators Used: 5

Remarks:  
N/A

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

SUSAN M BARGE

Signature and Printed Name

04/06/2016  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**

FEB 09 2016

**FDLE**

## AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: FL HIGHWAY PATROL  
Time of Inspection: 15:21

Date of Inspection: 02/01/2016

Serial Number: 80-003412  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 0014022 Exp: 07/14/2016	0.08g/210L Test (g/210L) Lot#: 0014022 Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 0014022 Exp: 02/24/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 0014022 Exp: 07/14/2016
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.078	0.197	0.079
0.000	0.050	0.078	0.198	0.079

Number of Simulators Used: 2

Remarks:  
NONE

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Gregory J. Landon*

\_\_\_\_\_  
GREGORY J. LANDON

Signature and Printed Name

02/01/2016  
Date

8780-003472

RECEIVED

# Florida Department of Law Enforcement Alcohol Testing Program

JAN 29 2016

FDLE

## AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: FL HIGHWAY PATROL  
Time of Inspection: 16:36

Date of Inspection: 01/28/2016

Serial Number: 80-003412  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 201502A Exp: 07/14/2017	0.05g/210L Test (g/210L) Lot#: 201502B Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#: 201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 16014060A2 Exp: 07/08/2016
0.000	0.049	0.079	0.197	0.078
0.000	0.049	0.078	0.198	0.079
0.000	0.049	0.078	0.196	0.078

Number of Simulators Used: 5

Remarks:  
NONE

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name GREGORY S. LAMBERT

01/28/2016  
Date