

RECEIVED

Florida Department of Law Enforcement  
Alcohol Testing Program

MAY 24 2016

FDLE  
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BEACH CO SO  
Time of Inspection: 11:30

Date of Inspection: 04/15/2016

Serial Number: 80-006029  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:13314080A3 Exp: 06/05/2016
0.000	0.049	0.080	0.202	0.081
0.000	0.050	0.080	0.202	0.080
0.000	0.050	0.080	0.202	0.081

Number of Simulators Used: 5

Remarks:

Time-Date changed. A F / M A: TO MUCH M/A

NOTE: ACCIDENTALLY TYPED OVER EXCEPTION MESSAGE. - J#6467

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*J#6467*  
Signature and Printed Name JASON C KARLECKE

04/15/2016  
Date