

Florida Department of Law Enforcement Alcohol Testing Program

FEB 05 2016

FDLE
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE CITY PD
Time of Inspection: 08:49

Date of Inspection: 02/04/2016

Serial Number: 80-006221
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | | No |
| Alcohol Free Subject Test: 0.000 | | No |
| Mouth Alcohol Test: Slope Not Met | | No |
| Interferent Detect Test: Interferent Detect | | No |
| Diagnostic Check (Post-Inspection): OK | | No |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: Exp: | 0.08g/210L Test (g/210L) Lot#: Exp: | 0.20g/210L Test (g/210L) Lot#: Exp: | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp: |
|----------------------------|--|--|--|---|
| | | | | |
| | | | | |
| | | | | |

Number of Simulators Used: _____

Remarks:
SKIPPED AI TO OPERATE INSTRUMENT

*N/A AI not conducted
Compliance not determined*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JENNIFER H JACOBS
Signature and Printed Name

02/04/2016
Date