

RECEIVED

MAR 10 2016

Florida Department of Law Enforcement
Alcohol Testing Program

COPY

FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ASTATULA PD

Time of Inspection: 10:17

Date of Inspection: 03/03/2016

Serial Number: 80-006222

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201411K Exp: 11/25/2016	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201408B Exp: 08/12/2016	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:09014080A1 Exp: 05/01/2016
0.000	0.046	0.075	0.190	0.080
0.000	0.046	0.075	0.191	0.080
0.000	0.046	0.076	0.192	0.080

Number of Simulators Used: 5

Remarks:

The readings have been passing, but rather low.

Can you check these out?

THANKS

Tim

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

TIMOTHY A PALINSKI

Signature and Printed Name

03/03/2016

Date