

# Florida Department of Law Enforcement

## Alcohol Testing Program

**RECEIVED**  
AUG 01 2016  
FDLE  
Alcohol Testing Program

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Hillsborough County SO  
Time of Inspection: 0900

Date of Inspection: 07/25/2016

Serial Number: 80-006385  
Software: 8100.27

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		
Alcohol Free Subject Test: 0.000		
Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect		
Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_

Remarks: Instrument passed start-up diagnostics but then displayed ERROR CODE #17" and immediately went to "STANDY MODE". Inspection cannot be performed prior to removing from service because of this condition. Instrument will be sent to repair facility.

Compliance could not be determined. This report manually generated.

The above instrument complies (\_\_\_\_) does not comply (\_\_\_\_) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
Roger G. Skipper #246109

Signature and Printed Name

07/25/2016  
Date

*Reviewed  
7-30-16  
MST, J.F. [Signature]*