

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ESCAMBIA COUNTY SO  
Time of Inspection: 05:49

Date of Inspection: 03/27/2016

Serial Number: 80-006435  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:03415080A2 Exp: 03/05/2017
0.000	0.043 / 0.046	0.078	0.201	0.000 / 0.080
0.000	0.043 / 0.046	0.078	0.200	0.000 / 0.080
0.000	0.044 / 0.046	0.078	0.200	0.080 / 0.080


Number of Simulators Used: 5

**Remarks:**

A F / M A: Purge Fail. 05: Control Outside Tolerance / FAULTY TUBE ON SIMULATOR. 08: Control Outside Tolerance / TUBE WASN'T CONNECTED.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
 \_\_\_\_\_  
 Signature and Printed Name

03/27/2016  
 Date