

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

MAY 31 2016
BK
FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ESCAMBIA COUNTY SO
Time of Inspection: 07:53

Date of Inspection: 05/26/2016

Serial Number: 80-006436
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

ogm


Number of Simulators Used: _____

Remarks:
SKIPPED AI TO OPERATE INSTRUMENT; NOT A COMPLIANCE CHECK

N/A (circled)

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PHILLIP W SUBER

Signature and Printed Name

05/26/2016
Date

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

MAY 31 2016
BK FDLE

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ESCAMBIA COUNTY SO
Time of Inspection: 07:53

Date of Inspection: 05/26/2016

Serial Number: 80-006436
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Wgm

Number of Simulators Used: _____

Remarks:

SKIPPED AI TO OPERATE INSTRUMENT; NOT A COMPLIANCE CHECK

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Phillip W Suber

PHILLIP W SUBER

Signature and Printed Name

05/26/2016
Date