

# Florida Department of Law Enforcement

## Alcohol Testing Program

**RECEIVED**

JAN 20 2016

FDLE

### AGENCY INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: HILLSBOROUGH CO SO  
 Time of Inspection: 09:53

Date of Inspection: 01/20/2016

Serial Number: 80-006567  
 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_

Remarks:  
 BYPASSED AI TO OPERATE INSTRUMENT. NOT A COMPLIANCE CHECK.

✓

N/A <sup>DOB</sup>

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Danielle M Bell*

DANIELLE M BELL

Signature and Printed Name

01/20/2016  
Date