

INSTRUMENT#
80-000788

Platt, Shayla

From: Platt, Shayla
Sent: Wednesday, October 12, 2016 8:47 AM
To: 'snowd@stluciesheriff.com'
Cc: 'MMarshall@alcoholtest.com'
Subject: FW: 80-000788
Attachments: 339506.pdf

Good Morning,

Please see the attached repair estimate for the instrument listed above. Please contact CMI and advise them how your agency would like to proceed with repairs.

Thanks,

Shayla

From: Marshall, Marianna [<mailto:MMarshall@alcoholtest.com>]
Sent: Tuesday, October 11, 2016 3:58 PM
To: Platt, Shayla
Subject: 80-000788

Good afternoon Shayla,

The instrument above was sent in for repair by David Caglione with St. Lucie PD. I have been unable to reach him to provide the estimate (attached). Could you possibly pass it along to him if you have his contact information?

Thanks so much,

Marianna Marshall
Service Coordinator
CMI Inc.
316 East 9th St.
Owensboro, KY 42303
(270) 685-6395 – Desk
(270) 685-6260 - Fax

MMarshall@alcoholtest.com

CMI, Inc.
 316 East Ninth Street
 Owensboro, KY 42303
 1 866 835 0690
 FAX (270) 685 6268

REPAIR ESTIMATE
 DATE 10/04/16
 PO# FDLE

PAGE 1

FROM Gary

CUSTOMER 349810
 ST LUCIE, COUNTY OF
 FT. PIERCE, FL
 CONTACT BILL SIMON
 PHONE 772-462-3359

S/N 80-000788 TECH GTW
 DESC UNIT ASSY, I8000
 SWO 339506

PRODUCT	DESCRIPTION	QTY	PRICE	EXT PRICE
150010	SPEAKER, 8 OHMS, .5W	1	3.900	3.90
103073	RES NET, SMD, 10P ARRAY, 8@6.8K	1	.250	.25
110184	CAP, .47UF, 400V, 10%, POLYESTER	1	1.250	1.25
R4	LABOR INTOX-1/2 HR INCREMENTS	5	42.500	212.50
R2	INTOX-CAL/FINAL TEST (IR)	1	79.000	79.00
R9	INTOX-CALIBRATION ADJUST (IR)	1	42.500	42.50
P1	PACKING & HANDLING - CMI	2	7.500	15.00

TOTAL LABOR 334.00
 TOTAL PARTS 5.40
 TOTAL PACKING/HANDLING 15.00
 TOTAL FREIGHT 15.00
 TOTAL TAX

TOTAL ESTIMATE 369.40

NOTE: AN EVALUATION FEE WILL APPLY TO ESTIMATES THAT ARE NOT REPAIRED.

PLEASE SIGN AND FAX TO: (270) 685 6268

APPROVED BY: _____

DATE: _____