

Bell, Danielle

From: Bell, Danielle
Sent: Monday, August 01, 2016 11:14 AM
To: 'Taylor, Keith'
Subject: RE: July Inspection

Thank you for the update.

Four things:

1. When keying in remarks in the remarks section, please do not backspace removing the "Correct Caps" type comments that are entered by the instrument. Simply start typing after those comments.
2. As far as I see, the first inspection (completed on July 31, 2106) was in compliance. Is there a reason you repeated the inspection?
3. The next inspection (completed August 1, 2016) has a notation of AF / MA. I take it you had to repeat the alcohol free / mouth alcohol portion of that inspection as well. Could you please make a notation on that Form 40 as to why there was a repeated AF/MA then initial and date it and send me a copy?
4. While these two inspections technically satisfy the inspections for July and August, I suggest (in the interest of smooth court proceedings) that you still do a normal monthly inspection closer to your usual date. (Maybe before midnight the night it would be "due".) *smiles

Thanks,

Danielle M. Bell
Department Inspector
Florida Department of Law Enforcement
Alcohol Testing Program
(850) 617-1286 (Office)
(850) 274-2757 (Cell)
(850) 921-3787 (Fax)

-----Original Message-----

From: Taylor, Keith [<mailto:TaylorK@ci.mount-dora.fl.us>]
Sent: Sunday, July 31, 2016 11:23 PM
To: Bell, Danielle
Subject: July Inspection

For July's inspection, I had another snafu with the M/A alcohol that was correct on the re-test. Also, my on duty radio triggered an RFI. Again this was corrected on the re-test.

CONFIDENTIALITY NOTICE: This e-mail message, including any attachments, is for the sole use of the intended recipient(s) and may contain confidential and privileged information. Any unauthorized review, use, disclosure, or distribution, is prohibited. If you are not the intended recipient, contact the sender by replying to the e-mail and destroy all copies of the original message.

PUBLIC RECORDS NOTICE: Under Florida law, e-mail addresses are public records. If you do not want your e-mail address released in response to a public records request, do not send electronic mail to this entity. Instead, contact this office by phone or in writing.

Bell, Danielle

From: Taylor, Keith <TaylorK@ci.mount-dora.fl.us>
Sent: Saturday, April 23, 2016 10:29 AM
To: Bell, Danielle
Subject: April Intox

Follow Up Flag: Follow Up
Due By: Wednesday, April 27, 2016 4:50 PM
Flag Status: Flagged

Hi there. I had a reading on this months AF/MA test. I presume I used too much mouth alcohol. I applied a small drop using an eye dropper on the tip of my tongue, and received a reading of .15. I do not understand how I can get a reading like this from such a small drop of mount alcohol. To me, it doesn't make since. You would think after ten years, I would have this mastered.

CONFIDENTIALITY NOTICE: This e-mail message, including any attachments, is for the sole use of the intended recipient(s) and may contain confidential and privileged information. Any unauthorized review, use, disclosure, or distribution, is prohibited. If you are not the intended recipient, contact the sender by replying to the e-mail and destroy all copies of the original message.

PUBLIC RECORDS NOTICE: Under Florida law, e-mail addresses are public records. If you do not want your e-mail address released in response to a public records request, do not send electronic mail to this entity. Instead, contact this office by phone or in writing.