

**Platt, Shayla**

---

**From:** Platt, Shayla  
**Sent:** Wednesday, September 07, 2016 9:39 AM  
**To:** 'Dustin.Worth@mysanibel.com'  
**Subject:** RE: Agency Instrument

Good Morning Dustin,

The internal battery needs to be replaced if the instrument is not maintaining the correct date and time. We are able to replace the battery here at FDLE.

Thanks,

Shayla

---

**From:** Alcohol Testing Program Mailbox  
**Sent:** Wednesday, September 07, 2016 6:17 AM  
**To:** Platt, Shayla  
**Subject:** FW: Agency Instrument

*Shayla, this is for you.*

*Barbara*

---

**From:** Dustin L. Worth [<mailto:Dustin.Worth@mysanibel.com>]  
**Sent:** Thursday, September 01, 2016 7:36 PM  
**To:** Alcohol Testing Program Mailbox  
**Subject:** Agency Instrument

I performed my August inspection on 08/29/2016 and had to reset the date and time. I did not realize the date had reset to the year 2000. I have attached the amended inspection with the corrected date.

Also, when conducting a breath test for a DUI arrest tonight, I had to adjust the date and time again. After the breath test, the instrument went into disabled mode due to the year 2000 issue. I am in the process of conducting another inspection to put it back into service.

Is the date/time issue an internal battery problem, and is that an issue FDLE can fix, or do I need to ship the instrument to my vendor, Enforcement Electronics?

Thank you,

Sgt. Dustin Worth  
Sanibel Police Department  
800 Dunlop Road  
Sanibel, FL 33957  
239-472-3111