

## Bell, Danielle

---

**From:** Jason Engel <EngelJ@CityofNewPortRichey.org>  
**Sent:** Thursday, October 06, 2016 8:22 PM  
**To:** Bell, Danielle  
**Subject:** Intoxilizer fail temp reg diagnostic

**Follow Up Flag:** Follow up  
**Flag Status:** Completed

Hello,

Per our conversation I attempted to conduct another inspection. The instrument again failed the temp reg diagnostic before the inspection. I would like to clarify before I send it out that it goes to you and not to Enforcement Electronics. Please advise.

Thank you,

Jason Engel  
Sergeant/Patrol Division  
New Port Richey Police  
6739 Adams Street  
New Port Richey, FL 34652  
727-841-4550 ext 124  
fax 727-816-1132

This correspondence is from a law enforcement agency and may contain confidential, secret, classified, and/or privileged communications. Unauthorized use of this information may subject you to criminal and civil penalties. This document is the property of the New Port Richey Police Department and may not be disseminated without authorization. As such this document may be exempt from public record as defined in Florida State Statute Chapter 119. It is the responsibility of the recipient of this correspondence to ensure applicable Laws, Statutes, Rules and Regulations are followed.

**\*\*\*\*\* NOTICE: This electronic transmission is confidential and intended only for the person(s) to whom it is addressed. If you have received this transmission in error, please notify the sender by return e-mail and destroy this message in its entirety (including all attachments). \*\*\*\*\***

**Bell, Danielle**

---

**From:** Jason Engel <EngelJ@CityofNewPortRichey.org>  
**Sent:** Friday, October 07, 2016 2:50 PM  
**To:** Bell, Danielle  
**Subject:** RE: Intoxilizer fail temp reg diagnostic

I did the upload last night. I will ship it out asap.

Thank you,

---

**From:** Bell, Danielle [<mailto:DanielleBell@fdle.state.fl.us>]  
**Sent:** Friday, October 07, 2016 8:06 AM  
**To:** Jason Engel  
**Subject:** RE: Intoxilizer fail temp reg diagnostic

Yes, it would come straight to me. I can determine which aspect may be causing the temp reg fail and possibly avoid extra costs to the agency. Also, please login to menu level 2 and go to comms transfer and upload the instrument prior to shipping.

Thanks,

*Danielle M. Bell*

Department Inspector  
Florida Department of Law Enforcement  
Alcohol Testing Program  
(850) 617-1286 (Office)  
(850) 274-2757 (Cell)  
(850) 921-3787 (Fax)

---

**From:** Jason Engel [<mailto:EngelJ@CityofNewPortRichey.org>]  
**Sent:** Thursday, October 06, 2016 8:22 PM  
**To:** Bell, Danielle  
**Subject:** Intoxilizer fail temp reg diagnostic

Hello,

Per our conversation I attempted to conduct another inspection. The instrument again failed the temp reg diagnostic before the inspection. I would like to clarify before I send it out that it goes to you and not to Enforcement Electronics. Please advise.

Thank you,

Jason Engel  
Sergeant/Patrol Division  
New Port Richey Police  
6739 Adams Street  
New Port Richey, FL 34652  
727-841-4550 ext 124  
fax 727-816-1132

This correspondence is from a law enforcement agency and may contain confidential, secret, classified, and/or privileged communications. Unauthorized use of this information may subject you to criminal and civil penalties. This document is the property of the New Port Richey Police Department and may not be disseminated without authorization. As such this document may be exempt from public record as defined in Florida State Statute Chapter 119. It is the responsibility of the recipient of this correspondence to ensure applicable Laws, Statutes, Rules and Regulations are followed.

**\*\*\*\*\* NOTICE: This electronic transmission is confidential and intended only for the person(s) to whom it is addressed. If you have received this transmission in error, please notify the sender by return e-mail and destroy this message in its entirety (including all attachments). \*\*\*\*\***