Bell, Danielle

From:

Ray Weber < RWeber@hainescitypd.com>

Sent:

Thursday, July 28, 2016 7:04 AM

To:

Bell, Danielle

Subject:

RE: 80-001151

Great. Thank you so very much.

IVE: 00-00113.

Ray

FDLE Alcohol Testing Program

RECEIVED

From: Bell, Danielle [mailto:DanielleBell@fdle.state.fl.us]

Sent: Thursday, July 28, 2016 7:02 AM

To: Ray Weber < RWeber@hainescitypd.com >

Subject: RE: 80-001151

No big deal. I can reprint the Form 40 from our end and send you a copy to write your explanation on, sign, date and send back. Any time you are unsure or if you end up having to repeat the inspection, just shoot me an email or call and often I'll get back with you pretty quickly. I should be in the office by 9 and I'll send the Form 40 when I get there.

Thanks, Danielle

Sent from my Verizon Wireless 4G LTE smartphone

----- Original message -----

From: Ray Weber < RWeber@hainescitypd.com >

Date: 07/28/2016 6:56 AM (GMT-05:00)

To: "Bell, Danielle" < <u>DanielleBell@fdle.state.fl.us</u>> Cc: Kevin Hager < <u>KHager@hainescitypd.com</u>>

Subject: RE: 80-001151

Ms. Bell, I received the form that you sent me. I apologize for the mess up on the first attempt of the inspection for June 2016. When I went to conduct the inspection, I accidentally pressed the #2 key twice instead of the Esc key which in turn caused a big mess for me and I was unable to abort the test. In order to quickly go through the prompts, I kept hitting the enter key so I can get back to the point where I was able to do the inspection correctly. I did not print the FDLE/ATP Form 40 out on the first inspection because I did not realize that it was also going to be uploaded along with the correct inspection. After doing the inspection correctly, I did print that form and sent it in. We went online to attempt to print out the incorrect inspection from the FDLE website, however the printout is not the FDLE/ATP Form 40 as requested. Can I still use the electronic printout "Inspection Test Electronic Data" from the FDLE website? If not, then are you able to print the Form 40 and email that to me? I really appreciate your feedback on how I can correct this. Thank you.

Cpl. Ray M. Weber II, #896

Haines City Police Department.

(863) 421-3636. Ext. 2234 (Office)

(863) 307-2550

(Cell)

From: Bell, Danielle [mailto:DanielleBell@fdle.state.fl.us]

Sent: Monday, July 25, 2016 3:55 PM

To: Ray Weber < RWeber@hainescitypd.com >

Subject: 80-001151 Importance: High

Please see attached.

Danielle M. Bell

Department Inspector

Florida Department of Law Enforcement

Alcohol Testing Program

(850) 617-1286 (Office)

(850) 274-2757 (Cell)

(850) 921-3787 (Fax)

Bell, Danielle

From:

Bell, Danielle

Sent:

Monday, July 25, 2016 3:55 PM

To:

Ray Weber

Subject:

80-001151

Attachments:

Weber AI Data Review.pdf

Importance:

High

Please see attached.

Danielle M. Bell

Department Inspector Florida Department of Law Enforcement Alcohol Testing Program (850) 617-1286 (Office) (850) 274-2757 (Cell) (850) 921-3787 (Fax)



Agency: Haines City P.D.

AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Instrument Serial Number: 80-001151

AGENCY INSPECTION DATA REVIEW Agency Inspector: Ray Weber Date of Inspection: 06/27/2016 Time of Inspection: 06:20 Agency Inspection Discrepancy: Untimely/Not Received Erroneous Information Incomplete Ø Procedural Other Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded. _g/210L □Alcohol Reference Solution □Dry Gas Standard is □Incorrect □Expired. FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report - Intoxilyzer 8000. The REASON for repeating the following test(s); OR the D Possible Cause and Corrective Action Taken on the following test(s) was not recorded: Alcohol Free Subject Test Mouth Alcohol Test

Alcohol Free Test Interferent Detect Test 0.05 g/210L Test 0.08 g/210L Test 0.20 g/210L Test 0.08 g/210L Dry Gas Standard Test FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector." The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC. The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use. The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use. The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC. \boxtimes Other: OTHER ELECTRONIC DATA REVIEW Comments: Login Records Date: Please write in the comments section on the original form 40 why the inspection was not completed. Please also mark through the X indicating compliance (with a single line) and Cylinder Change Records mark the "non-compliance" space. Please initial and date all changes and send a copy to Date: Control Test Records me. Diagnostic Check Records CORRECTIVE ACTION Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by 07/30/2016 (Date). Provide a written explanation regarding the referenced item(s) to the Department Inspector by (Date). Upload the Agency Inspection(s). Remove the instrument from evidentiary use until otherwise directed by the Department. No action required Other: 7/21/2016 Signature of Alcohol Testing Program Staff Member Date