

80-001345 *QMG*

EUSTIS P.D.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001345
01/18/2016
Software: 8100.27

RECEIVED

JAN 19 2016

**FDLE
Alcohol Testing Program**

DIAGNOSTICS

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK