

**Bell, Danielle**

80-001057  
80-001865  
80-003945  
80-005810

**From:** Franz, Gordon <Gordon.Franz@lakelandgov.net>  
**Sent:** Tuesday, December 06, 2016 3:30 PM  
**To:** Stephanie Cornejo (scorejo@sao10.com); skaylor@sao10.com; Bell, Danielle; Trulson, Tom; Cari Daniels; scorejo@sao10.com; wdennis@sao10.com; Miranda, Ivette; Taylor, Sammy; Harper, Eric V; Mumbauer, Chad; Kellner, Michael; Suttle, Cory; Quackenbush, Bret  
**Cc:** Franz, Gordon  
**Subject:** LPD's Intoxilyzer Inspections for December 2016  
**Attachments:** LPD's Intoxilyzer Inspections for December 2016.pdf

To whom it may concern,

Please see the attached files for the Lakeland Police Department's monthly Intoxilyzer Inspections for December 2016.

1. Intoxilyzer Serial # 80-001057, monthly inspection completed (In-Service).
2. Intoxilyzer Serial # 80-001865, monthly inspection completed (In-Service).
3. Intoxilyzer Serial # 80-003945, monthly inspection completed (In-Service).
4. Intoxilyzer Serial # 80-005810, monthly inspection completed (In-Service).

*Gordon L. Franz, MS*

Red Light Camera Program | Agency Inspector  
Lakeland Police Department | Special Ops Section | City of Lakeland  
219 North Massachusetts Avenue  
Lakeland, Florida 33801  
Office 863-834-2560 Cell 813-416-5497 | [Gordon.Franz@lakelandgov.net](mailto:Gordon.Franz@lakelandgov.net)  
*Working together to achieve an exceptional quality of life*

---

PUBLIC RECORDS NOTICE:

All e-mail sent to and received from the City of Lakeland, Florida, including e-mail addresses and content, are subject to the provisions of the Florida Public Records Law, Florida Statute Chapter 119, and may be subject to disclosure.

80-001057  
80-001865  
80-003945  
80-005810

**Bell, Danielle**

---

**From:** Franz, Gordon <Gordon.Franz@lakelandgov.net>  
**Sent:** Wednesday, November 30, 2016 8:50 AM  
**To:** Stephanie Cornejo (scorejo@sao10.com); skaylor@sao10.com; Bell, Danielle; Trulson, Tom; Cari Daniels; scorejo@sao10.com; wdennis@sao10.com; Miranda, Ivette; Taylor, Sammy; Harper, Eric V; Mumbauer, Chad; Kellner, Michael; Suttle, Cory; Quackenbush, Bret  
**Cc:** Franz, Gordon  
**Subject:** LPD's Intoxilyzer Inspections for November 2016  
**Attachments:** LPD Intoxilyzer Inspections November 2016.pdf

To whom it may concern,

Please see the attached files for the Lakeland Police Department's monthly Intoxilyzer Inspections for November 2016.

1. Intoxilyzer Serial # 80-001057, monthly inspection completed (In-Service).
2. Intoxilyzer Serial # 80-001865, shipped to FDLE for the annual inspection.
3. Intoxilyzer Serial # 80-003945, monthly inspection completed (In-Service).
4. Intoxilyzer Serial # 80-005810, shipped to FDLE for the annual inspection.

*Gordon L. Franz, MS*

Red Light Camera Program | Agency Inspector  
Lakeland Police Department | Special Ops Section | City of Lakeland  
219 North Massachusetts Avenue  
Lakeland, Florida 33801  
Office 863-834-2560 Cell 813-416-5497 | [Gordon.Franz@lakelandgov.net](mailto:Gordon.Franz@lakelandgov.net)  
*Working together to achieve an exceptional quality of life*

---

**PUBLIC RECORDS NOTICE:**

All e-mail sent to and received from the City of Lakeland, Florida, including e-mail addresses and content, are subject to the provisions of the Florida Public Records Law, Florida Statute Chapter 119, and may be subject to disclosure.

**Bell, Danielle**

---

**From:** Franz, Gordon <Gordon.Franz@lakelandgov.net>  
**Sent:** Tuesday, October 25, 2016 9:40 AM  
**To:** Stephanie Cornejo (scorejo@sao10.com); skaylor@sao10.com; Bell, Danielle; Trulson, Tom; Cari Daniels; scorejo@sao10.com; wdennis@sao10.com; Miranda, Ivette; Taylor, Sammy; Harper, Eric V; Mumbauer, Chad; Kellner, Michael; Suttle, Cory; Quackenbush, Bret  
**Subject:** LPD's Intoxilyzer Inspections for October 2016  
**Attachments:** LPD's Intoxilyzer Inspections for October 2016.pdf

To whom it may concern,

Please see the attached files for the Lakeland Police Department's monthly Intoxilyzer Inspections for October 2016.

1. Intoxilyzer Serial # 80-001057, monthly inspection completed (In-Service).
2. Intoxilyzer Serial # 80-001865, currently at Enforcement Electronics for maintenance.
3. Intoxilyzer Serial # 80-003945, monthly inspection completed (In-Service).
4. Intoxilyzer Serial # 80-005810, currently at Enforcement Electronics for maintenance.

*Gordon L. Franz, MS*

Red Light Camera Program | Agency Inspector  
Lakeland Police Department | Special Ops Section | City of Lakeland  
219 North Massachusetts Avenue  
Lakeland, Florida 33801  
Office 863-834-2560 Cell 813-416-5497 | [Gordon.Franz@lakelandgov.net](mailto:Gordon.Franz@lakelandgov.net)  
*Working together to achieve an exceptional quality of life*

---

**PUBLIC RECORDS NOTICE:**

All e-mail sent to and received from the City of Lakeland, Florida, including e-mail addresses and content, are subject to the provisions of the Florida Public Records Law, Florida Statute Chapter 119, and may be subject to disclosure.

80-001057  
80-001865  
80-003945  
80-005810

**Bell, Danielle**

---

**From:** Franz, Gordon <Gordon.Franz@lakelandgov.net>  
**Sent:** Tuesday, October 11, 2016 3:25 PM  
**To:** Stephanie Cornejo (scorejo@sao10.com); skaylor@sao10.com; Bell, Danielle; Trulson, Tom; Cari Daniels; scorejo@sao10.com; wdennis@sao10.com; Miranda, Ivette; Taylor, Sammy; Harper, Eric V; Mumbauer, Chad; Kellner, Michael; Suttle, Cory; Quackenbush, Bret  
**Subject:** Lakeland Police Departments Intoxilyzer Inspections for September 2016  
**Attachments:** LPD Intoxilyzer Inspections for September 2016.pdf

To whom it may concern,

Please see the attached files for the Lakeland Police Department's monthly Intoxilyzer Inspections for September 2016.

1. Intoxilyzer Serial # 80-001057, the annual inspection was completed by FDLE and was received back during the week of 09/26/2016.
2. Intoxilyzer Serial # 80-001865, was outside tolerance during monthly inspections, removed from service, taken to Enforcement Electronic for maintenance.
3. Intoxilyzer Serial # 80-003945, monthly inspection completed (In-Service).
4. Intoxilyzer Serial # 80-005810, was shipped to Enforcement Electronic by FDLE and is scheduled for its annual preventive maintenance.

*Gordon L. Franz, MS*

Red Light Camera Program | Agency Inspector  
Lakeland Police Department | Special Ops Section | City of Lakeland  
219 North Massachusetts Avenue  
Lakeland, Florida 33801  
Office 863-834-2560 Cell 813-416-5497 | [Gordon.Franz@lakelandgov.net](mailto:Gordon.Franz@lakelandgov.net)  
*Working together to achieve an exceptional quality of life*

---

**PUBLIC RECORDS NOTICE:**

All e-mail sent to and received from the City of Lakeland, Florida, including e-mail addresses and content, are subject to the provisions of the Florida Public Records Law, Florida Statute Chapter 119, and may be subject to disclosure.

80-001057  
80-003945  
80-005810  
80-001865

**Bell, Danielle**

---

**From:** Franz, Gordon <Gordon.Franz@lakelandgov.net>  
**Sent:** Monday, August 22, 2016 3:02 PM  
**To:** Stephanie Cornejo (sornejo@sao10.com); skaylor@sao10.com; Bell, Danielle; Trulson, Tom; Cari Daniels; sornejo@sao10.com; skaylor@sao10.com; wdennis@sao10.com; Miranda, Ivette; Taylor, Sammy; Harper, Eric V; Mumbauer, Chad; Kellner, Michael; Suttle, Cory; Quackenbush, Bret  
**Subject:** Lakeland Police Departments Intoxilyzer Inspections for August 2016  
**Attachments:** LPD's Intoxilyzer Room Inspections for August 2016.docx.pdf  
**Follow Up Flag:** Follow up  
**Flag Status:** Flagged

To whom it may concern,

Please see the attached file for the Lakeland Police Department's monthly Intoxilyzer Inspections for August 2016.

Intoxilyzer # 80-001057 was delivered to Enforcement Electronics on 06/13/2016 for an annual PM.

Intoxilyzer # 80-003945 is enroute to FDLE for the annual inspection.

Intoxilyzer # 80-005810 is scheduled to be shipped directly to FDLE during the week of August 22 – 26, 2016.

*Gordon L. Franz, MS*

Red Light Camera Program | Agency Inspector  
Lakeland Police Department | Special Ops Section | City of Lakeland  
219 North Massachusetts Avenue  
Lakeland, Florida 33801  
Office 863-834-2560 Cell 813-416-5497 | [Gordon.Franz@lakelandgov.net](mailto:Gordon.Franz@lakelandgov.net)  
*Working together to achieve an exceptional quality of life*

---

**PUBLIC RECORDS NOTICE:**

All e-mail sent to and received from the City of Lakeland, Florida, including e-mail addresses and content, are subject to the provisions of the Florida Public Records Law, Florida Statute Chapter 119, and may be subject to disclosure.

**Bell, Danielle**

---

**From:** Franz, Gordon <Gordon.Franz@lakelandgov.net>  
**Sent:** Wednesday, October 19, 2016 6:43 AM  
**To:** Bell, Danielle  
**Cc:** Mumbauer, Chad  
**Subject:** Intoxilyzer # 80-001865  
**Attachments:** Updated Intoxilyzer Inspection for Sept 2016, Serial # 80-001865.pdf

Good Morning Mrs. Bell,

Attached you will find the updated Form 40 for Intoxilyzer # 80-001865 for September 2016. Please let me know if there is anything else I am required to do.

Respectfully,

*Gordon L. Franz, MS*

Red Light Camera Program | Agency Inspector  
Lakeland Police Department | Special Ops Section | City of Lakeland  
219 North Massachusetts Avenue  
Lakeland, Florida 33801  
Office 863-834-2560 Cell 813-416-5497 | [Gordon.Franz@lakelandgov.net](mailto:Gordon.Franz@lakelandgov.net)  
*Working together to achieve an exceptional quality of life*

---

**PUBLIC RECORDS NOTICE:**

All e-mail sent to and received from the City of Lakeland, Florida, including e-mail addresses and content, are subject to the provisions of the Florida Public Records Law, Florida Statute Chapter 119, and may be subject to disclosure.

**Bell, Danielle**

---

**From:** Bell, Danielle  
**Sent:** Monday, February 15, 2016 3:16 PM  
**To:** 'Franz, Gordon'; Mumbauer, Chad  
**Cc:** Suttle, Cory  
**Subject:** RE: Intoxilyzer issue

Not a problem. We shouldn't have that issue again, however, if you make a habit of checking the recall screen after uploading, that should let you know if there is another issue. If it happens again, please let me know.

Thanks,

*Danielle M. Bell*

Department Inspector  
 Florida Department of Law Enforcement  
 Alcohol Testing Program  
 (850) 617-1286 (Office)  
 (850) 274-2757 (Cell)  
 (850) 921-3787 (Fax)

---

**From:** Franz, Gordon [<mailto:Gordon.Franz@lakelandgov.net>]  
**Sent:** Monday, February 15, 2016 2:49 PM  
**To:** Bell, Danielle; Mumbauer, Chad  
**Cc:** Suttle, Cory  
**Subject:** RE: Intoxilyzer issue

Good Afternoon Mrs. Bell,

After I spoke with you this morning, I was able to transmit the January's inspection report for instrument 890-001865 (held in memory) without difficulties. I then completed the February's inspection for instruments 80-001057, 80-005810 and 80-001865. The reports were subsequently uploaded and I received the message "Upload OK". I also verified that no reports were held in memory. The transmission of the monthly reports appear to be working fine. I will keep you posted if I have any further glitches. Thanks for all your assistance!

*Gordon L. Franz, MS*

Red Light Camera Program | Agency Inspector  
 Lakeland Police Department | Special Ops Section | City of Lakeland  
 219 North Massachusetts Avenue  
 Lakeland, Florida 33801  
 Office 863-834-2560 Cell 813-416-5497 | [Gordon.Franz@lakelandgov.net](mailto:Gordon.Franz@lakelandgov.net)  
*Working together to achieve an exceptional quality of life*

---

**From:** Bell, Danielle [<mailto:DanielleBell@fdle.state.fl.us>]  
**Sent:** Monday, February 15, 2016 8:22 AM

**To:** Franz, Gordon; Mumbauer, Chad  
**Subject:** RE: Intoxilyzer issue

Do we have an update on this?

*Danielle M. Bell*

Department Inspector  
Florida Department of Law Enforcement  
Alcohol Testing Program  
(850) 617-1286 (Office)  
(850) 274-2757 (Cell)  
(850) 921-3787 (Fax)

---

**From:** Bell, Danielle  
**Sent:** Wednesday, February 10, 2016 9:31 AM  
**To:** 'Franz, Gordon'  
**Cc:** Mumbauer, Chad  
**Subject:** RE: Intoxilyzer issue

Good morning,

It does appear that instrument #1865 uploaded just fine and I can see the inspection. If you would, at your convenience, please plug in the instrument, turn it on, and give me a call once it warms up. Sometimes it takes a few long seconds for that message to clear once it appears, but I wouldn't think it would be there since you uploaded same day. I would like to check a few things before we do anything else.

Thanks,

*Danielle M. Bell*

Department Inspector  
Florida Department of Law Enforcement  
Alcohol Testing Program  
(850) 617-1286 (Office)  
(850) 274-2757 (Cell)  
(850) 921-3787 (Fax)

---

**From:** Franz, Gordon [<mailto:Gordon.Franz@lakelandgov.net>]  
**Sent:** Wednesday, February 10, 2016 6:49 AM  
**To:** Bell, Danielle  
**Cc:** Mumbauer, Chad  
**Subject:** FW: Intoxilyzer issue

Good Morning Mrs. Danielle Bell,

Below is an email I received from Officer Cory Suttle, one of our Breath Test Operators. The Intoxilyzer inspections were completed on January 25, 2016 and all 4 instruments passed. Each inspection result was uploaded via the "phone lines" and usually takes several attempts to make a connection and upload the files. Each instrument indicated that the upload was completed "OK".

Did you receive the electronic version of our Agency Inspection Report (Form 40) for instrument # 80-001865 dated January 25, 2016?



I have instructed our BTO's to use one of the other instruments until I can verify that the report was successfully received. I will be completing the February's inspections on Monday, February 15, 2016.

*Gordon L. Franz, MS*

Red Light Camera Program | Agency Inspector

Lakeland Police Department | Special Ops Section | City of Lakeland

219 North Massachusetts Avenue

Lakeland, Florida 33801

Office 863-834-2560 Cell 813-416-5497 | [Gordon.Franz@lakelandgov.net](mailto:Gordon.Franz@lakelandgov.net)

*Working together to achieve an exceptional quality of life*

---

**From:** Suttle, Cory

**Sent:** Monday, February 08, 2016 7:37 AM

**To:** Franz, Gordon

**Cc:** Kellner, Michael; Mumbauer, Chad

**Subject:** Intoxilyzer issue

I did a breath test this morning and got an error message on the instrument with the serial number ending with 1865. The message was "Disabled Mode AI Update Expired". I have moved it into the cabinet and hooked up the other instrument that was in storage. Both remaining instruments seem to be working fine and I did perform a test with one.

Officer Cory Suttle #279

DUI Enforcement Unit

Lakeland Police Department

863-834-2551

---

**PUBLIC RECORDS NOTICE:**

All e-mail sent to and received from the City of Lakeland, Florida, including e-mail addresses and content, are subject to the provisions of the Florida Public Records Law, Florida Statute Chapter 119, and may be subject to disclosure.