

RECEIVED

JUL 01 2016

**Severance, Mark D.**

**Subject:** FW: Intoxilyzer Failure

**FDLE**  
**Alcohol Testing Program**

Attempted to conduct an agency inspection – still received a persistent **\*\*Ambient Fail\*\*** multiple times. As discussed on the phone, I will send you the instrument for repair/inspection. It will need to be downloaded, since I did not complete the inspection or upload for the month of June.

**From:** Severance, Mark D.  
**Sent:** Tuesday, June 21, 2016 4:03 PM  
**To:** 'jakeshanahan@fdle.state.fl.us' <jakeshanahan@fdle.state.fl.us>  
**Subject:** Intoxilyzer Failure

Jake,

I have one of our instruments currently out of service for an consistent **\*\*Ambient Fail\*\*** . As part of my trying to troubleshoot the issue, I performed a T-Control Testing D-Dry Control Test and still got **\*\*Ambient Fail\*\***. What shall I do now? Can you assist me with troubleshooting? I have not completed the monthly agency inspection for June as of this time.

**Sgt Mark D. Severance #1215**

New Smyrna Beach Police Department  
246 Industrial Park Avenue  
New Smyrna Beach, Florida 32168  
Email: [mseverance@cityofnsb.com](mailto:mseverance@cityofnsb.com)  
Phone: (386) 402-7788  
Cell: (386) 690-1004

**PLEASE NOTE:** Florida has a very broad public records law. E-mail communications to or from City of New Smyrna Beach officials and employees are public records and are available upon request. If you have received this message in error, please notify us immediately by replying to this message and delete it from your computer. Transmission of sensitive information such as but not limited to social security numbers, credit card information, medical records and criminal justice information ( outlined in the [FBI CJIS Policy](#) ) via email is prohibited.

*\* Please consider the negative environmental impacts associated with printing email \**

**VISIT OUR WEBSITE FOR MORE INFORMATION ABOUT THE CITY OF NEW SMYRNA BEACH FLORIDA**  
[www.cityofnsb.com](http://www.cityofnsb.com)

RECEIVED

JUL 01 2016

FDLE  
Alcohol Testing Program

NEW SMYRNA BEACH PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-002231  
06/17/2016  
Software: 8100.27

-----  
DRY CONTROL TEST

Test	g/210L	Time
Air Blank	AMB*	10:47
Air Blank	AMB*	10:47

\*Ambient Fail

\_\_\_\_\_  
Operator's Signature

Do Not

\* Use \*

Soft. Severance  
6-17-16

RECEIVED

JUL 01 2016

FDLE  
Alcohol Testing Program

NEW SMYRNA BEACH PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-002231  
06/17/2016  
Software: 8100.27

-----  
DRY CONTROL TEST

Test	g/210L	Time
Air Blank	AMB*	10:40
Air Blank	AMB*	10:41

\*Ambient Fail

\_\_\_\_\_  
Operator's Signature