




City of Crestview
POLICE DEPARTMENT

201 Stillwell Boulevard – Crestview, Florida 32539
Tony R. Taylor, Chief of Police



Administration
850.682.3544
850.682.2080 fax

Communications center
850.682.2055
Fax 850.689.2175

Agency Inspector

To: Florida Department of Law Enforcement Alcohol Testing Program

Attention: Phillip Suber

From: Andrew Schneider (Agency Inspector)

Re: March 2016 Inspection for 80-005058

Mr. Suber,

During the month of March 2016, the Crestview Police Department was unable to conduct an Agency Inspection on Instrument # 80-005058 due to a shortage of simulator solutions. The previous passing inspection was conducted on February 29, 2016.

On April 01, 2016, beginning at midnight, the Instrument became disabled and was out of service as a result of an automatic instrument protocol and it was no longer available for evidentiary use.

On April 04, 2016, solutions were obtained and an Agency Inspection was conducted in accordance to the requirements of the Florida Administrative Codes. The results of this testing was within passing tolerances and the instrument was returned in service for evidentiary sampling.

This letter is to inform you that records will indicate that there was no inspection conducted in the month of March. However, an inspection was conducted at the first available opportunity when the solutions for testing were received.

Should you have any questions, please feel free to contact me at your earliest convenience.

Respectfully,

A handwritten signature in black ink, appearing to read "Andrew Schneider".

Andrew Schneider
Agency Inspector
Crestview Police Department

Duty, Honor, Courage, Integrity