

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

DEC 08 2016 *vec*

FDLE

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

BT

Agency: FDLE

Serial Number: 80-000224

Time of Inspection: 13:56

Date of Inspection: 12/07/2016

Software: 8100.27

| Check or Test | YES | NO | Check or Test | YES | NO |
|---|-----|----|--|-----|----|
| Diagnostic Check (Pre-Inspection): OK | Yes | | Date and/or Time Adjusted | | No |
| Minimum Sample Volume Check: OK | Yes | | Barometric Pressure Sensor Check: OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | | Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | | Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017 | 0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018 | 0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG619605 Exp: 07/14/2018 |
|----------------------------|---|---|---|---|
| 0.000 | 0.048 | 0.079 | 0.201 | 0.077 |
| 0.000 | 0.048 | 0.080 | 0.203 | 0.078 |
| 0.000 | 0.048 | 0.080 | 0.202 | 0.077 |
| 0.000 | 0.049 | 0.081 | 0.204 | 0.077 |
| 0.000 | 0.049 | 0.080 | 0.203 | 0.078 |
| 0.000 | 0.049 | 0.080 | 0.203 | 0.078 |
| 0.000 | 0.048 | 0.080 | 0.203 | 0.078 |
| 0.000 | 0.049 | 0.080 | 0.204 | 0.078 |
| 0.000 | 0.049 | 0.078 | 0.204 | 0.078 |
| 0.000 | 0.049 | 0.077 | 0.204 | 0.077 |

| | | | | |
|---------------------|--------|--------|--------|--------|
| Standard Deviations | 0.0005 | 0.0011 | 0.0009 | 0.0005 |
|---------------------|--------|--------|--------|--------|

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0007 Number of Simulators Used: 5

Remarks:

SP

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Danielle M Bell

DANIELLE M BELL

Signature and Printed Name

12/07/2016
Date

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

DEC 08 2016 *vec*

FDLE

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

BT

Agency: FDLE

Serial Number: 80-000224

Time of Inspection: 13:56

Date of Inspection: 12/07/2016

Software: 8100.27

| Check or Test | YES | NO | Check or Test | YES | NO |
|---|-----|----|--|-----|----|
| Diagnostic Check (Pre-Inspection): OK | Yes | | Date and/or Time Adjusted | | No |
| Minimum Sample Volume Check: OK | Yes | | Barometric Pressure Sensor Check: OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | | Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | | Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017 | 0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018 | 0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG619605 Exp: 07/14/2018 |
|----------------------------|---|---|---|---|
| 0.000 | 0.048 | 0.079 | 0.201 | 0.077 |
| 0.000 | 0.048 | 0.080 | 0.203 | 0.078 |
| 0.000 | 0.048 | 0.080 | 0.202 | 0.077 |
| 0.000 | 0.049 | 0.081 | 0.204 | 0.077 |
| 0.000 | 0.049 | 0.080 | 0.203 | 0.078 |
| 0.000 | 0.049 | 0.080 | 0.203 | 0.078 |
| 0.000 | 0.048 | 0.080 | 0.203 | 0.078 |
| 0.000 | 0.049 | 0.080 | 0.204 | 0.078 |
| 0.000 | 0.049 | 0.078 | 0.204 | 0.078 |
| 0.000 | 0.049 | 0.077 | 0.204 | 0.077 |

| | | | | |
|---------------------|--------|--------|--------|--------|
| Standard Deviations | 0.0005 | 0.0011 | 0.0009 | 0.0005 |
|---------------------|--------|--------|--------|--------|

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0007 Number of Simulators Used: 5

Remarks:

SP

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Danielle M Bell

DANIELLE M BELL

Signature and Printed Name

12/07/2016
Date