

# Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED *VLS*

MAY 24 2016

FDLE

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000 Alcohol Testing Program

Agency: ALACHUA COUNTY SO  
Time of Inspection: 13:33

Date of Inspection: 05/21/2016

Serial Number: 80-000737  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG600504 Exp: 01/05/2018
0.000	0.050	0.080	0.200	0.080
0.000	0.050	0.079	0.200	0.079
0.000	0.050	0.080	0.201	0.079
0.000	0.050	0.080	0.201	0.079
0.000	0.051	0.081	0.201	0.079
0.000	0.050	0.081	0.201	0.078
0.000	0.050	0.080	0.200	0.079
0.000	0.051	0.080	0.200	0.079
0.000	0.050	0.080	0.200	0.079
0.000	0.051	0.080	0.200	0.079
0.000	0.050	0.080	0.200	0.079
0.000	0.051	0.080	0.200	0.078

Standard Deviations	0.0004	0.0005	0.0005	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

*(Handwritten mark)*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Patrick J. Murphy*  
\_\_\_\_\_  
Signature and Printed Name

05/21/2016  
Date