

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
JUL 06 2016
FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MARION COUNTY SO
Time of Inspection: 14:00

Date of Inspection: 07/01/2016

Serial Number: 80-000752
Software: 8100.27

| Check or Test | YES | NO | Check or Test | YES | NO |
|--|-----|----|---|-----|----|
| Diagnostic Check (Pre-Inspection): OK | Yes | | Date and/or Time Adjusted | | No |
| Minimum Sample Volume Check: OK | Yes | | Barometric Pressure Sensor Check: OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | | Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | | Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017 | 0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018 | 0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG605301 Exp: 02/22/2018 |
|----------------------------------|--|--|--|---|
| 0.000 | 0.048 | 0.079 | 0.197 | 0.080 |
| 0.000 | 0.049 | 0.080 | 0.199 | 0.079 |
| 0.000 | 0.048 | 0.079 | 0.200 | 0.080 |
| 0.000 | 0.049 | 0.079 | 0.200 | 0.079 |
| 0.000 | 0.049 | 0.079 | 0.200 | 0.080 |
| 0.000 | 0.049 | 0.080 | 0.199 | 0.080 |
| 0.000 | 0.049 | 0.079 | 0.199 | 0.080 |
| 0.000 | 0.049 | 0.079 | 0.199 | 0.080 |
| 0.000 | 0.049 | 0.079 | 0.199 | 0.080 |
| 0.000 | 0.049 | 0.079 | 0.199 | 0.080 |
| 0.000 | 0.050 | 0.079 | 0.199 | 0.079 |

| | | | | |
|---------------------|--------|--------|--------|--------|
| Standard Deviations | 0.0005 | 0.0004 | 0.0008 | 0.0004 |
|---------------------|--------|--------|--------|--------|

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Patrick J Murphy

PATRICK J MURPHY

Signature and Printed Name

07/01/2016
Date