

RECEIVED *1/2*

Florida Department of Law Enforcement Alcohol Testing Program

FEB 22 2016

FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 15:00

Date of Inspection: 02/19/2016

Serial Number: 80-000759
Software: 8100.27

| Check or Test | YES | NO | Check or Test | YES | NO |
|---|-----|----|--|-----|----|
| Diagnostic Check (Pre-Inspection): OK | Yes | | Date and/or Time Adjusted | | No |
| Minimum Sample Volume Check: OK | Yes | | Barometric Pressure Sensor Check: OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | | Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | | Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017 | 0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017 | 0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG511701 Exp: 04/27/2017 |
|----------------------------|---|---|---|---|
| 0.000 | 0.049 | 0.076 | 0.202 | 0.078 |
| 0.000 | 0.049 | 0.078 | 0.203 | 0.077 |
| 0.000 | 0.049 | 0.078 | 0.203 | 0.076 |
| 0.000 | 0.049 | 0.078 | 0.204 | 0.077 |
| 0.000 | 0.049 | 0.078 | 0.203 | 0.077 |
| 0.000 | 0.049 | 0.078 | 0.204 | 0.077 |
| 0.000 | 0.049 | 0.078 | 0.203 | 0.077 |
| 0.000 | 0.049 | 0.079 | 0.204 | 0.077 |
| 0.000 | 0.050 | 0.078 | 0.203 | 0.077 |
| 0.000 | 0.050 | 0.079 | 0.203 | 0.077 |

| | | | | |
|---------------------|--------|--------|--------|--------|
| Standard Deviations | 0.0004 | 0.0008 | 0.0006 | 0.0004 |
|---------------------|--------|--------|--------|--------|

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

PHILLIP W SUBER

Signature and Printed Name

02/19/2016
Date