

RECEIVED

## Florida Department of Law Enforcement DEC 09 2016

## Alcohol Testing Program

FDLE  
Alcohol Testing ProgramDEPARTMENT INSPECTION REPORT - INTOXILYZER 8000 *BCC*Agency: NAS JACKSONVILLE PD  
Time of Inspection: 10:48

Date of Inspection: 12/09/2016

Serial Number: 80-000778  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626605 Exp: 09/22/2018
0.000	0.082 / 0.049	0.079	0.201	0.079
0.000	0.081 / 0.050	0.080	0.202	0.079
0.000	0.081 / 0.049	0.080	0.202	0.079
0.000	0.082 / 0.049	0.080	0.203	0.079
0.000	0.082 / 0.050	0.080	0.201	0.080
0.000	0.084 / 0.049	0.081	0.204	0.080
0.000	/ 0.050	0.081	0.201	0.079
0.000	/ 0.050	0.082	0.204	0.080
0.000	/ 0.051	0.080	0.202	0.081
0.000	/ 0.051	0.082	0.202	0.080

Standard Deviations	/ 0.0007	0.0009	0.0011	0.0006

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0008 Number of Simulators Used: 5

## Remarks:

05: Purge Fail. INCORRECT SIM. REPEATED OK

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*JLS*  
*BLK*  
JAKE L SHANAHAN

Signature and Printed Name

12/09/2016  
Date