

RECEIVED

JUN 10 2016 ^{BLC}

Florida Department of Law Enforcement Alcohol Testing Program

FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MADISON COUNTY SO
Time of Inspection: 14:19

Date of Inspection: 06/09/2016

Serial Number: 80-000779
Software: 8100.27

| Check or Test | YES | NO | Check or Test | YES | NO |
|---|-----|----|--|-----|----|
| Diagnostic Check (Pre-Inspection): OK | Yes | | Date and/or Time Adjusted | | No |
| Minimum Sample Volume Check: OK | Yes | | Barometric Pressure Sensor Check: OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | | Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | | Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017 | 0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018 | 0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG600504 Exp: 01/05/2018 |
|----------------------------|---|---|---|---|
| 0.000 | 0.049 | 0.079 | 0.201 | 0.081 |
| 0.000 | 0.050 | 0.079 | 0.204 | 0.081 |
| 0.000 | 0.049 | 0.078 | 0.204 | 0.081 |
| 0.000 | 0.050 | 0.078 | 0.203 | 0.080 |
| 0.000 | 0.050 | 0.078 | 0.203 | 0.081 |
| 0.000 | 0.050 | 0.077 | 0.203 | 0.081 |
| 0.000 | 0.049 | 0.080 | 0.200 | 0.081 |
| 0.000 | 0.051 | 0.079 | 0.202 | 0.081 |
| 0.000 | 0.050 | 0.078 | 0.201 | 0.081 |
| 0.000 | 0.050 | 0.079 | 0.202 | 0.081 |

PSM

| | | | | |
|---------------------|--------|--------|--------|--------|
| Standard Deviations | 0.0006 | 0.0008 | 0.0013 | 0.0003 |
|---------------------|--------|--------|--------|--------|

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0007 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

PHILLIP W SUBER

Signature and Printed Name

06/09/2016
Date

Stability Tests - Madison CSO # 80-000779 6/9/16

ASK

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
06/09/2016
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 10:57 |
| Control Test | 0.050 | 10:58 |
| Air Blank | 0.000 | 10:59 |
| Control Test | 0.049 | 10:59 |
| Air Blank | 0.000 | 11:00 |
| Control Test | 0.051 | 11:01 |
| Air Blank | 0.000 | 11:01 |
| Control Test Stats | | |
| Average | 0.0500 | |
| Std Dev | 0.0010 | |
| Rel Std Dev(%) | 2.0000 | |

PMS

Operator's Signature

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
06/09/2016
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 11:03 |
| Control Test | 0.080 | 11:04 |
| Air Blank | 0.000 | 11:04 |
| Control Test | 0.080 | 11:05 |
| Air Blank | 0.000 | 11:05 |
| Control Test | 0.081 | 11:06 |
| Air Blank | 0.000 | 11:06 |
| Control Test Stats | | |
| Average | 0.0803 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 0.7187 | |

PMS

Operator's Signature

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
06/09/2016
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 11:09 |
| Control Test | 0.204 | 11:09 |
| Air Blank | 0.000 | 11:10 |
| Control Test | 0.204 | 11:11 |
| Air Blank | 0.000 | 11:11 |
| Control Test | 0.204 | 11:12 |
| Air Blank | 0.000 | 11:12 |
| Control Test Stats | | |
| Average | 0.2040 | |
| Std Dev | 0.0000 | |
| Rel Std Dev(%) | 0.0000 | |

PMS

Operator's Signature

MADISON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
06/09/2016
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 11:15 |
| Control Test | 0.081 | 11:15 |
| Air Blank | 0.000 | 11:16 |
| Control Test | 0.080 | 11:16 |
| Air Blank | 0.000 | 11:17 |
| Control Test | 0.080 | 11:17 |
| Air Blank | 0.000 | 11:17 |
| Control Test Stats | | |
| Average | 0.0803 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 0.7187 | |

PMS

Operator's Signature

DES