

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

SEP 16 2016 *BCE*

FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

SK

Agency: GULF BREEZE PD
Time of Inspection: 14:34

Date of Inspection: 09/15/2016

Serial Number: 80-000787
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG612405 Exp: 05/03/2018
0.000	0.048	0.078	0.194	0.081
0.000	0.048	0.079	0.197	0.080
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.079	0.197	0.080
0.000	0.049	0.079	0.196	0.080
0.000	0.049	0.079	0.197	0.080
0.000	0.048	0.079	0.197	0.079
0.000	0.049	0.078	0.197	0.079
0.000	0.049	0.078	0.196	0.079
0.000	0.049	0.078	0.196	0.080

QJM


Standard Deviations	0.0004	0.0005	0.0009	0.0006
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PHILLIP W SUBER

Signature and Printed Name

09/15/2016
Date

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

JUN 10 2016

FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF BREEZE PD
Time of Inspection: 14:06

Date of Inspection: 06/09/2016

Serial Number: 80-000787
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG600504 Exp: 01/05/2018
0.000	0.049	0.078	0.200	0.080
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.078	0.202	0.079
0.000	0.050	0.079	0.202	0.079
0.000	0.050	0.079	0.203	0.079
0.000	0.049	0.079	0.203	0.079
0.000	0.049	0.079	0.203	0.079
0.000	0.049	0.079	0.203	0.079
0.000	0.049	0.078	0.203	0.079
0.000	0.050	0.079	0.203	0.079


Standard Deviations	0.0004	0.0004	0.0010	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PHILLIP W SUBER

Signature and Printed Name

06/09/2016
Date

Stability Tests - Gulf Breeze PD #50-000787 6/4/16

AK

GULF BREEZE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000787
06/09/2016
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	10:03
Control Test	0.050	10:04
Air Blank	0.000	10:04
Control Test	0.049	10:05
Air Blank	0.000	10:05
Control Test	0.050	10:06
Air Blank	0.000	10:07
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

(Handwritten initials)

(Handwritten signature)
Operator's Signature

GULF BREEZE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000787
06/09/2016
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	10:09
Control Test	0.078	10:10
Air Blank	0.000	10:11
Control Test	0.079	10:11
Air Blank	0.000	10:12
Control Test	0.078	10:13
Air Blank	0.000	10:13
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

(Handwritten signature)
Operator's Signature

GULF BREEZE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000787
06/09/2016
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	10:15
Control Test	0.199	10:16
Air Blank	0.000	10:16
Control Test	0.198	10:17
Air Blank	0.000	10:17
Control Test	0.199	10:18
Air Blank	0.000	10:19
Control Test Stats		
Average	0.1987	
Std Dev	0.0006	
Rel Std Dev(%)	0.2906	

(Handwritten signature)
Operator's Signature

GULF BREEZE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000787
06/09/2016
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	09:58
Control Test	0.080	09:58
Air Blank	0.000	09:59
Control Test	0.080	09:59
Air Blank	0.000	10:00
Control Test	0.080	10:00
Air Blank	0.000	10:00
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DAV

(Handwritten signature)
Operator's Signature