

RECEIVED

Florida Department of Law Enforcement

Alcohol Testing Program

SEP 07 2016 *vac*

FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MONROE COUNTY SO
Time of Inspection: 12:41

Date of Inspection: 08/22/2016

Serial Number: 80-000870
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG600504 Exp: 01/05/2018
0.000	0.051	0.080	0.206	0.081
0.000	0.051	0.081	0.207	0.081
0.000	0.050	0.082	0.208	0.081
0.000	0.051	0.081	0.207	0.081
0.000	0.051	0.080	0.207	0.080
0.000	0.051	0.081	0.207	0.081
0.000	0.051	0.081	0.207	0.080
0.000	0.051	0.081	0.206	0.081
0.000	0.051	0.080	0.206	0.081
0.000	0.051	0.081	0.206	0.081
Standard Deviations	0.0003	0.0006	0.0006	0.0004

egm

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David Reyes Reyes Rivera

Signature and Printed Name

DAVID E REYES-RIVERA

08/22/2016
Date