## Florida Department of Law Enforcement DEC 0 8 2016 Alcohol Testing Program

FOLE Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

BLC
- Jan

Agency	:	rwc-raim	Beach
Time o	£	Inspection:	

Interferent Detect

Date of Inspection: 12 4 16 Software: 8100.27

(Post-Inspection): OK

Serial Number: 80-000903

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK			Date and/or Time Adjusted		
Minimum Sample Volume			Barometric Pressure Sensor		
Check: OK	To a second seco		Check: OK		
Alcohol Free Subject			Mouth Alcohol Test:		
Test: 0.000			Slope Not Met		
Interferent Detect Test:			Diagnostic Check		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

	Standard Deviations	
1	average Standard Deviation of 0.05, 0.08 ar	nd 0.20 g/210L Tests: Number of Simulators Used:
1	Remarks: INSTRUMENT FAILET	DSP DIAGNOSTIC CHECK PREVENTING
	IT FROM GOING	INTO READY MODE. UNABLE TO
	COMPLETE INSPECTI	

NIA COMPUANCE NO.
The above instrument complies ( ) does not comply ( ) with Chapter 11D-8, FAC.
I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.
Signature and Printed Name
12/4/11e

FDLE/ATP Form 41 -Revised August 2005