

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

OCT 18 2016

FDLE

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PINECREST PD
Time of Inspection: 14:55

Date of Inspection: 09/29/2016

Serial Number: 80-000909
Software: 8100.27

| Check or Test | YES | NO | Check or Test | YES | NO |
|---|-----|----|--|-----|----|
| Diagnostic Check (Pre-Inspection): OK | Yes | | Date and/or Time Adjusted | | No |
| Minimum Sample Volume Check: OK | Yes | | Barometric Pressure Sensor Check: OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | | Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | | Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017 | 0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018 | 0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG600504 Exp: 01/05/2018 |
|----------------------------|---|---|---|---|
| 0.000 | 0.050 | 0.080 | 0.204 | 0.081 |
| 0.000 | 0.050 | 0.081 | 0.205 | 0.080 |
| 0.000 | 0.051 | 0.081 | 0.205 | 0.080 |
| 0.000 | 0.051 | 0.081 | 0.204 | 0.080 |
| 0.000 | 0.051 | 0.081 | 0.204 | 0.080 |
| 0.000 | 0.051 | 0.081 | 0.203 | 0.080 |
| 0.000 | 0.052 | 0.081 | 0.203 | 0.080 |
| 0.000 | 0.051 | 0.081 | 0.203 | 0.081 |
| 0.000 | 0.051 | 0.081 | 0.203 | 0.080 |
| 0.000 | 0.051 | 0.081 | 0.204 | 0.080 |

| | | | | |
|---------------------|--------|--------|--------|--------|
| Standard Deviations | 0.0005 | 0.0003 | 0.0007 | 0.0004 |
|---------------------|--------|--------|--------|--------|

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DAVID E REYES-RIVERA

Signature and Printed Name

09/29/2016
Date

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
APR 12 2016
FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PINECREST PD
Time of Inspection: 10:08

Date of Inspection: 03/28/2016

Serial Number: 80-000909
Software: 8100.27

| Check or Test | YES | NO | Check or Test | YES | NO |
|---|-----|----|--|-----|----|
| Diagnostic Check (Pre-Inspection): OK | Yes | | Date and/or Time Adjusted | | No |
| Minimum Sample Volume Check: OK | Yes | | Barometric Pressure Sensor Check: OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | | Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | | Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017 | 0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017 | 0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG519701 Exp: 07/16/2017 |
|----------------------------|---|---|---|---|
| 0.000 | 0.051 | 0.079 | 0.197 | 0.080 |
| 0.000 | 0.051 | 0.080 | 0.197 | 0.080 |
| 0.000 | 0.050 | 0.080 | 0.198 | 0.080 |
| 0.000 | 0.051 | 0.080 | 0.199 | 0.079 |
| 0.000 | 0.051 | 0.080 | 0.199 | 0.079 |
| 0.000 | 0.051 | 0.079 | 0.199 | 0.079 |
| 0.000 | 0.051 | 0.079 | 0.199 | 0.079 |
| 0.000 | 0.051 | 0.080 | 0.199 | 0.079 |
| 0.000 | 0.051 | 0.079 | 0.199 | 0.079 |
| 0.000 | 0.051 | 0.080 | 0.199 | 0.079 |

RM

| | | | | |
|---------------------|--------|--------|--------|--------|
| Standard Deviations | 0.0003 | 0.0005 | 0.0008 | 0.0004 |
|---------------------|--------|--------|--------|--------|

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David Reyes-Rivera

DAVID E REYES-RIVERA

Signature and Printed Name

03/28/2016
Date