

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

OCT 28 2016 *Yes*

FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

BK

Agency: QUINCY PD
Time of Inspection: 11:38

Date of Inspection: 10/28/2016

Serial Number: 80-000965
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626605 Exp: 09/22/2018
0.000	0.048	0.079	0.197	0.080
0.000	0.049	0.080	0.197	0.080
0.000	0.049	0.080	0.198	0.080
0.000	0.049	0.080	0.198	0.080
0.000	0.048	0.081	0.198	0.080
0.000	0.049	0.081	0.197	0.079
0.000	0.048	0.081	0.197	0.080
0.000	0.049	0.081	0.197	0.080
0.000	0.049	0.081	0.197	0.080
0.000	0.048	0.081	0.197	0.079

Standard Deviations	0.0005	0.0007	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

QUB

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Patrick J Murphy

PATRICK J MURPHY

Signature and Printed Name

10/28/2016
Date

RECEIVED

Florida Department of Law Enforcement Alcohol Testing Program

OCT 28 2016 *Yes*

FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

BK

Agency: QUINCY PD
Time of Inspection: 11:38

Date of Inspection: 10/28/2016

Serial Number: 80-000965
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626605 Exp: 09/22/2018
0.000	0.048	0.079	0.197	0.080
0.000	0.049	0.080	0.197	0.080
0.000	0.049	0.080	0.198	0.080
0.000	0.049	0.080	0.198	0.080
0.000	0.048	0.081	0.198	0.080
0.000	0.049	0.081	0.197	0.079
0.000	0.048	0.081	0.197	0.080
0.000	0.049	0.081	0.197	0.080
0.000	0.049	0.081	0.197	0.080
0.000	0.048	0.081	0.197	0.079

Standard Deviations	0.0005	0.0007	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

QMB

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Patrick J Murphy

PATRICK J MURPHY

Signature and Printed Name

10/28/2016
Date