

Florida Department of Law Enforcement Alcohol Testing Program

NOV 09 2016 ✓

FDLE
Alcohol Testing Program
BK

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SEMINOLE COUNTY S.O.
Time of Inspection: 10:49

Date of Inspection: 11/07/2016

Serial Number: 80-000999
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG626605 Exp: 07/22/2018
0.000	0.051	0.082	0.201	0.000 / 0.078
0.000	0.051	0.082	0.203	0.000 / 0.078
0.000	0.051	0.082	0.202	0.000 / 0.078
0.000	0.051	0.082	0.202	0.077 / 0.078
0.000	0.051	0.083	0.202	0.077 / 0.078
0.000	0.051	0.082	0.202	0.078 / 0.078
0.000	0.051	0.082	0.202	0.078 / 0.078
0.000	0.051	0.083	0.202	0.077 / 0.078
0.000	0.052	0.082	0.201	0.078 / 0.078
0.000	0.051	0.082	0.202	0.077 / 0.078

Standard Deviations	0.0003	0.0004	0.0005	0.0374 / 0.0000
---------------------	--------	--------	--------	-----------------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5


Remarks:

08: Control Outside Tolerance DID NOT HAVE DG CONNECTED

*Retested after connected.
In compliance ✓*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JAKE L SHANAHAN

Signature and Printed Name

11/07/2016
Date