

Florida Department of Law Enforcement Alcohol Testing Program

JUN 15 2016

FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 12:23

Date of Inspection: 06/07/2016

Serial Number: 80-001047
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG600504 Exp: 01/05/2018
0.000	0.051	0.081	0.204	0.079
0.000	0.051	0.081	0.204	0.079
0.000	0.051	0.081	0.204	0.080
0.000	0.051	0.081	0.204	0.079
0.000	0.052	0.081	0.203	0.080
0.000	0.051	0.081	0.203	0.079
0.000	0.052	0.081	0.204	0.079
0.000	0.052	0.081	0.203	0.079
0.000	0.051	0.082	0.204	0.079
0.000	0.052	0.081	0.204	0.079
Standard Deviations	0.0005	0.0003	0.0004	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David Reyes Rivera

Signature and Printed Name

06/07/2016
Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-001047	Broward County Sheriff's Office	06/07/2016	<i>AKL</i>

AK

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>

BROWARD COUNTY SO
Intoxilyzer - Alcotest Analyzer
Model 8000
06/07/2016
SN 80-001047
Software: 8100.27

Test	g/210L	Time
Alr Blank	0.000	09:58
Control Test	0.050	09:58
Alr Blank	0.000	09:59
Control Test	0.051	10:00
Alr Blank	0.000	10:00
Control Test	0.051	10:01
Alr Blank	0.000	10:01
Control Test	0.051	10:02
Control Test Stats		
Average	0.0507	
Std Dev	0.0006	
Rel Std Dev(%)	1.1395	

AKL
Operator's Signature

BROWARD COUNTY SO
Intoxilyzer - Alcotest Analyzer
Model 8000
06/07/2016
SN 80-001047
Software: 8100.27

Test	g/210L	Time
Alr Blank	0.000	10:03
Control Test	0.082	10:03
Alr Blank	0.000	10:04
Control Test	0.081	10:04
Alr Blank	0.000	10:05
Control Test	0.081	10:06
Alr Blank	0.000	10:06
Control Test	0.081	10:06
Control Test Stats		
Average	0.0813	
Std Dev	0.0006	
Rel Std Dev(%)	0.7099	

AKL
Operator's Signature

BROWARD COUNTY SO
Intoxilyzer - Alcotest Analyzer
Model 8000
06/07/2016
SN 80-001047
Software: 8100.27

Test	g/210L	Time
Alr Blank	0.000	10:08
Control Test	0.203	10:09
Alr Blank	0.000	10:09
Control Test	0.203	10:10
Alr Blank	0.000	10:10
Control Test	0.203	10:11
Alr Blank	0.000	10:11
Control Test	0.203	10:12
Control Test Stats		
Average	0.2030	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

AKL
Operator's Signature

BROWARD COUNTY SO
Intoxilyzer - Alcotest Analyzer
Model 8000
06/07/2016
SN 80-001047
Software: 8100.27

Test	g/210L	Time
Alr Blank	0.000	10:14
Control Test	0.079	10:15
Alr Blank	0.000	10:15
Control Test	0.080	10:16
Alr Blank	0.000	10:16
Control Test	0.079	10:16
Alr Blank	0.000	10:17
Control Test	0.000	10:17
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

AKL
Operator's Signature

gym

RECEIVED ✓/PAC

Florida Department of Law Enforcement

Alcohol Testing Program

JAN 21 2016
FDLE Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO Serial Number: 80-001047
Time of Inspection: 10:19 Date of Inspection: 01/04/2016 Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG431502 Exp: 11/11/2016
0.000	0.050	0.081	0.202	0.079
0.000	0.051	0.080	0.203	0.079
0.000	0.050	0.080	0.204	0.079
0.000	0.051	0.080	0.203	0.079
0.000	0.051	0.081	0.204	0.079
0.000	0.051	0.080	0.204	0.079
0.000	0.051	0.080	0.204	0.079
0.000	0.051	0.081	0.204	0.079
0.000	0.051	0.080	0.204	0.079
0.000	0.050	0.080	0.204	0.079

Standard Deviations	0.0004	0.0004	0.0006	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David Reyes Rivera DAVID E REYES-RIVERA
Signature and Printed Name

01/04/2016
Date