

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED *Yes*

JUL 15 2016

FDLE

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000 *Alcohol Testing Program*

Agency: GULFPORT PD
Time of Inspection: 16:12

Date of Inspection: 07/14/2016

Serial Number: 80-001061
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG612405 Exp: 05/03/2018
0.000	0.047	0.077	0.197	0.085
0.000	0.048	0.078	0.196	0.084
0.000	0.048	0.078	0.196	0.083
0.000	0.048	0.078	0.199	0.084
0.000	0.048	0.078	0.199	0.083
0.000	0.049	0.077	0.198	0.083
0.000	0.049	0.078	0.197	0.083
0.000	0.048	0.079	0.198	0.083
0.000	0.048	0.078	0.197	0.082
0.000	0.048	0.078	0.197	0.083

Standard Deviations	0.0005	0.0005	0.0010	0.0008
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0007 Number of Simulators Used: 5

Remarks:

(Handwritten mark)

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

(Signature) DANIELLE M BELL
Signature and Printed Name

07/14/2016
Date