

RECEIVED

Florida Department of Law Enforcement JUL 27 2016

Alcohol Testing Program

FDLE Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: TARPON SPRINGS PD
Time of Inspection: 20:36

Date of Inspection: 07/26/2016

Serial Number: 80-001065
Software: 8100.27

Table with 6 columns: Check or Test, YES, NO, Check or Test, YES, NO. Rows include Diagnostic Check, Minimum Sample Volume, Alcohol Free Subject, and Interferent Detect Test.

Table with 5 columns: Alcohol Free Test (g/210L), 0.05g/210L Test, 0.08g/210L Test, 0.20g/210L Test, 0.08 g/210L Dry Gas Std Test. Multiple rows of numerical data.

Table with 5 columns: Standard Deviations, 0.0005, 0.0008, 0.0009, 0.0005

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

Handwritten mark/signature

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Handwritten signature

DANIELLE M BELL

Signature and Printed Name

07/26/2016
Date