

RECEIVED *off*

Florida Department of Law Enforcement FEB 25 2016 Alcohol Testing Program

FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO
Time of Inspection: 12:40

Date of Inspection: 02/25/2016

Serial Number: 80-001074
Software: 8100.27

| Check or Test | YES | NO | Check or Test | YES | NO |
|---|-----|----|--|-----|----|
| Diagnostic Check (Pre-Inspection): OK | Yes | | Date and/or Time Adjusted | | No |
| Minimum Sample Volume Check: OK | Yes | | Barometric Pressure Sensor Check: OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | | Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | | Diagnostic Check (Post-Inspection): OK | Yes | |

BK

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017 | 0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017 | 0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:09014080A1 Exp: 05/01/2016 |
|----------------------------|---|---|---|---|
| 0.000 | 0.050 | 0.080 | 0.201 | 0.080 |
| 0.000 | 0.051 | 0.080 | 0.201 | 0.079 |
| 0.000 | 0.051 | 0.080 | 0.201 | 0.079 |
| 0.000 | 0.051 | 0.080 | 0.202 | 0.079 |
| 0.000 | 0.051 | 0.080 | 0.201 | 0.079 |
| 0.000 | 0.051 | 0.081 | 0.202 | 0.079 |
| 0.000 | 0.051 | 0.080 | 0.201 | 0.079 |
| 0.000 | 0.051 | 0.080 | 0.201 | 0.079 |
| 0.000 | 0.051 | 0.080 | 0.201 | 0.079 |
| 0.000 | 0.051 | 0.080 | 0.201 | 0.079 |
| 0.000 | 0.051 | 0.081 | 0.201 | 0.079 |
| Standard Deviations | 0.0003 | 0.0004 | 0.0004 | 0.0003 |

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

PHS

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Danielle M Bell

DANIELLE M BELL

Signature and Printed Name

02/25/2016
Date