

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED
SEP 16 2016
FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: EGLIN AFB Serial Number: 80-001116
Time of Inspection: 12:18 Date of Inspection: 09/15/2016 Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.048	0.079		
0.000	0.049	0.080		
0.000	0.049	0.080		
0.000	0.049	0.079		
0.000	0.049	0.080		
0.000	0.049	0.079		
0.000	0.049	0.080		
0.000	0.049			
0.000	0.049			
0.000	0.049			

Standard Deviations	0.0003			
---------------------	--------	--	--	--


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: 5

Remarks:

08: RFI Detect UNKNOWN CAUSE; WILL RETEST. RETEST FAILED Non-compliance:RFI DURING 08 TEST. SENDIN G OUT FOR SERVICE..

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PHILLIP W SUBER

Signature and Printed Name

09/15/2016
Date