

Florida Department of Law Enforcement DEC 09 2016

Alcohol Testing Program

FDLE Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000 *VSC*

Agency: PUTNAM COUNTY SO
Time of Inspection: 11:08

Date of Inspection: 12/09/2016

Serial Number: 80-001129
Software: 8100.27

| Check or Test | YES | NO | Check or Test | YES | NO |
|---|-----|----|--|-----|----|
| Diagnostic Check (Pre-Inspection): OK | Yes | | Date and/or Time Adjusted | | No |
| Minimum Sample Volume Check: OK | Yes | | Barometric Pressure Sensor Check: OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | | Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | | Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:201603D Exp: 03/08/2018 | 0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018 | 0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG619605 Exp: 07/14/2018 |
|----------------------------|---|---|---|---|
| 0.000 | 0.049 | 0.081 | 0.198 | 0.000 / 0.081 |
| 0.000 | 0.049 | 0.081 | 0.198 | 0.080 / 0.081 |
| 0.000 | 0.050 | 0.081 | 0.199 | 0.081 / 0.081 |
| 0.000 | 0.049 | 0.081 | 0.198 | 0.080 / 0.081 |
| 0.000 | 0.050 | 0.081 | 0.198 | 0.080 / 0.080 |
| 0.000 | 0.050 | 0.081 | 0.199 | 0.080 / 0.081 |
| 0.000 | 0.050 | 0.082 | 0.198 | 0.080 / 0.081 |
| 0.000 | 0.050 | 0.082 | 0.198 | 0.080 / 0.081 |
| 0.000 | 0.050 | 0.081 | 0.198 | 0.080 / 0.081 |
| 0.000 | 0.050 | 0.081 | 0.199 | 0.080 / 0.081 |

| | | | | |
|---------------------|--------|--------|--------|-----------------|
| Standard Deviations | 0.0004 | 0.0004 | 0.0004 | 0.0253 / 0.0003 |
|---------------------|--------|--------|--------|-----------------|

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

08: Control Outside TolerancedRY GAS NOT CONNECTED REPEAT.

fk

[Signature]

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

[Signature]

JAKE L SHANAHAN

Signature and Printed Name

12/09/2016
Date