

RECEIVED *W*

Florida Department of Law Enforcement MAR 03 2016 Alcohol Testing Program FDLE Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE HELEN P.D. Serial Number: 80-001149
Time of Inspection: 11:08 Date of Inspection: 03/03/2016 Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

SK

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AGS11701 Exp: 04/27/2017
0.000	0.052	0.083	0.206	0.080
0.000	0.053	0.083	0.207	0.079
0.000	0.053	0.083	0.207	0.079
0.000	0.053	0.083	0.207	0.079
0.000	0.053	0.083	0.207	0.080
0.000	0.053	0.083	0.207	0.079
0.000	0.053	0.084	0.206	0.079
0.000	0.053	0.083	0.207	0.080
0.000	0.053	0.083	0.207	0.079
0.000	0.053	0.084	0.207	0.079

Standard Deviations	0.0003	0.0004	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.
I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JAKE L SHANAHAN
Signature and Printed Name

03/03/2016
Date

Florida Department of Law Enforcement

Alcohol Testing Program

FDLE
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE HELEN P.D.
Time of Inspection: 11:09

Date of Inspection: 12/15/2015

Serial Number: 80-001149
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AGS11701 Exp: 04/27/2017
0.000	0.052	0.083	0.207	0.078
0.000	0.052	0.083	0.207	0.078
0.000	0.053	0.083	0.207	0.078
0.000	0.053	0.083	0.207	0.078
0.000	0.053	0.083	0.207	0.078
0.000	0.053	0.084	0.208	0.078
0.000	0.053	0.084	0.208	0.078
0.000	0.053	0.084	0.208	0.078
0.000	0.053	0.083	0.208	0.079
0.000	0.053	0.084	0.208	0.078

Standard Deviations	0.0004	0.0005	0.0005	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

JMS

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JAKE L SHANAHAN

Signature and Printed Name

12/15/2015
Date