

# Florida Department of Law Enforcement Alcohol Testing Program

**RECEIVED**  
OCT 20 2016  
FDLE  
Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO  
Time of Inspection: 10:55

Date of Inspection: 10/20/2016

Serial Number: 80-001150  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201604C Exp: 04/05/2018	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG612405 Exp: 05/03/2018
0.000	0.047	0.078	0.197	0.079
0.000	0.048	0.079	0.197	0.080
0.000	0.048	0.079	0.198	0.080
0.000	0.049	0.079	0.198	0.080
0.000	0.048	0.080	0.198	0.080
0.000	0.049	0.080	0.198	0.080
0.000	0.049	0.079	0.198	0.080
0.000	0.049	0.079	0.198	0.080
0.000	0.049	0.080	0.198	0.080
0.000	0.049	0.079	0.199	0.080

Standard Deviations	0.0007	0.0006	0.0005	0.0003
---------------------	--------	--------	--------	--------


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:



The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
 \_\_\_\_\_  
 Signature and Printed Name SHAYLA D PLATT  
 \_\_\_\_\_  
 10/20/2016  
 Date

# Florida Department of Law Enforcement Alcohol Testing Program

OCT 20 2016

FDLE  
Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: Manatee County SO  
Time of Inspection:

Date of Inspection: 10/19/16

Serial Number: 80-001150  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK			Date and/or Time Adjusted		
Minimum Sample Volume Check: OK			Barometric Pressure Sensor Check: OK		
Alcohol Free Subject Test: 0.000			Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect			Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Standard Deviations				
---------------------	--	--	--	--

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: \_\_\_\_\_ Number of Simulators Used: \_\_\_\_\_

Remarks:

**During the final diagnostic check of my inspection I was plugging up the instrument to the printer and the electrical cord was not securely inserted into the instrument and came out of place resulting in a loss of power to the instrument. The inspection was compliant and had no issues prior to the loss of power.**

*N/A COMPLIANCE NOT DETERMINED*

The above instrument complies ( ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Shayla Platt*

Signature and Printed Name

10/19/16  
Date