

Florida Department of Law Enforcement Alcohol Testing Program

RECEIVED

APR 26 2016

FDLE

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: VOLUSIA COUNTY SO
Time of Inspection: 11:57

Date of Inspection: 04/26/2016

Serial Number: 80-001154
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201601F Exp: 01/26/2018	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG605301 Exp: 02/22/2018
0.000	0.049	0.077	0.202	0.079
0.000	0.050	0.078	0.203	0.080
0.000	0.051	0.075	0.203	0.079
0.000	0.050	0.075	0.202	0.079
0.000	0.050	0.075	0.202	0.079
0.000	0.051	0.079	0.202	0.079
0.000	0.050	0.079	0.202	0.079
0.000	0.051	0.079	0.202	0.079
0.000	0.050	0.080	0.202	0.079
0.000	0.050	0.075	0.202	0.079

Standard Deviations	0.0006	0.0020	0.0004	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0008 Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JAKE L SHANAHAN

Signature and Printed Name

04/26/2016
Date

RECEIVED

Florida Department of Law Enforcement

Alcohol Testing Program

APR 26 2016

FDLE

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: VOLUSIA COUNTY SO

Serial Number: 80-001154

Time of Inspection: 10:39

Date of Inspection: 03/08/2016

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201507A Exp: 07/14/2017	0.08g/210L Test (g/210L) Lot#:201502G Exp: 02/24/2017	0.20g/210L Test (g/210L) Lot#:201505A Exp: 05/12/2017	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG600504 Exp: 01/05/2018
0.000	0.052	0.081	0.206	0.080
0.000	0.053	0.083	0.206	0.079
0.000	0.053	0.083	0.207	0.079
0.000	0.052	0.083	0.207	0.080
0.000	0.053	0.083	0.207	0.080
0.000	0.053	0.083	0.208	0.079
0.000	0.053	0.084	0.207	0.080
0.000	0.053	0.084	0.207	0.079
0.000	0.053	0.084	0.207	0.079
0.000	0.054	0.084	0.207	0.080
0.000	0.053	0.084	0.207	0.080

Standard Deviations	0.0005	0.0009	0.0005	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JAKE L SHANAHAN

Signature and Printed Name

03/08/2016
Date